

**HUMAN SOLUBLE LECTIN -
LIKE OXIDIZED LOW-
DENSITY LIPOPROTEIN
RECEPTOR-1 (LOX-1)
ULTRASENSITIVE ELISA KIT**

FOR THE QUANTITATIVE DETERMINATION OF
HUMAN SOLUBLE LOX-1 CONCENTRATIONS IN CELL
CULTURE SUPERNATES, SERUM AND EDTA PLASMA



ALWAYS REFER TO LOT SPECIFIC PROTOCOL
PROVIDED WITH EACH KIT FOR
INSTRUCTIONS. PROTOCOL MUST BE
READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN
DIAGNOSTIC PROCEDURES.

PURCHASE INFORMATION:

THIS KIT IS FOR ONE TIME USE ONLY.

ELISA NAME	HUMAN SOLUBLE LOX-1 ULTRASENSITIVE ELISA
Catalog No.	SK00006-08
Lot No.	
Formulation	96 T
Standard range	7.8 - 500 pg/mL
Sensitivity	5 pg/mL
Sample Volume	100 µL
Dilution Factor	Direct Assay (Optimal dilutions should be determined by each laboratory for each application)
Sample Type	Serum, EDTA Plasma, Cell Culture Supernates
Specificity	Human sLOX-1
Calibration	Human sLOX-1 Rec. (NS0 derived)
Intra-assay Precision	4 - 6%
Inter-assay Precision	8 - 10%
Storage	2 – 8° C for 1 month. More information check page 2
This kit contains sufficient materials to run approximately 35 samples duplicated provided that assay is run according to protocol.	

Order Contact:

AVISCIERA BIOSCIENCE, INC.

2348 Walsh Ave., Suite C

Santa Clara, CA 95051

USA

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: Info@AvisceraBioscience.com

www.AvisceraBioscience.com

DESCRIPTION

This Human Soluble LOX-1 Ultrasensitive ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human soluble LOX-1 from cell culture supernates, serum and plasma in a sandwich ELISA format.

This immunoassay contains recombinant human soluble LOX-1 and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural soluble LOX-1 samples.

ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with a monoclonal antibody specific for human soluble LOX-1. The capture antibody can bind to the human soluble LOX-1 in the standard and samples. After washing the plate of any unbound substances, a biotinylated monoclonal antibody against human soluble LOX-1 is added to the wells. After another washing of the plate, Streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution (TMB) is added to the wells and color develops in direct proportion to the amount of human soluble LOX-1 bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Each laboratory must determine the optimal dilution factors for the samples being assayed.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
LOX-1 Microplate - 96 well polystyrene microplate (12 strips of 8 wells) coated with an antibody against human LOX-1.	006-08-01	1 plate
LOX-1 Standard – refer to lot specific of recombinant human soluble LOX-1 in a buffered protein base with preservative; lyophilized.	006-08-02	1 vial
Detection Antibody Concentrate – refer to lot specific, 10-fold concentrate of biotinylated antibody against human LOX-1 with preservative; lyophilized.	006-08-03	1 vial
Positive Control – one vial of recombinant human soluble LOX-1; lyophilized.	006-08-04	1 vial
Streptavidin-HRP Conjugate - 120 µL/vial, 100-fold concentrated solution of Streptavidin conjugate to HRP.	SAHRP	1 vial
Dilution Buffer – 40 mL of buffered protein based solution with preservative.	DB03	1 bottle
Antibody & HRP Diluent Solution – 25 mL of buffered protein based solution with preservative.	DB12	1 bottle
Wash Buffer – 50 mL of 10-fold concentrated buffered surfactant with preservative.	WB01	1 bottle
TMB Substrate Solution - 11 mL of TMB substrate solution.	TMB01	1 bottle
Stop Solution – 11 mL of 0.5M HCl.	S-STOP	1 bottle
Plate Sealer	EAPS	1
Plastic Pouch	P01	1

STORAGE

Unopened Kit: Store at 2 – 8° C for up to 1 month. For longer storage up to 10 months, unopened Standard, Positive Control, Detection Antibody Concentrate and Dilution Buffer should be stored at -20° C. Streptavidin-HRP Conjugate and TMB

Substrate Solution should be stored only at 2 – 8°C.
Do not use kit past expiration date.

Use polypropylene test tubes.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

SAMPLE COLLECTION AND STORAGE

Cell Culture Supernates - Remove particulates by centrifugation and assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Serum - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Plasma - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Optional: Use Aprotinin (enzyme inhibitor) for ALL sample collection to prevent sample degradation. 0.5 TIU per ml of sample solution.

SAMPLE PREPARATION

Plasma and serum samples DO NOT need to be diluted in this assay. However, if samples are higher than the 500 pg/mL maximum standard point, then a 2-fold dilution or greater might be needed. **Optimal dilutions should be determined by each laboratory for each application.**

REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of 1x Wash Buffer.

LOX-1 Standard - Reconstitute the LOX-1 standard with refer to lot specific of Dilution Buffer. Pipette 250 μL of Dilution Buffer into tubes #2 to #6. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The **625 pg/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 pg/mL).

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
stock	powder	Refer to lot	
# 1	Refer to lot	Refer to lot	500 pg/mL
# 2	250 μL of 1	250 μL	250 pg/mL
# 3	250 μL of 2	250 μL	125 pg/mL
# 4	250 μL of 3	250 μL	62.5 pg/mL
# 5	250 μL of 4	250 μL	31.25 pg/mL
# 6	250 μL of 5	250 μL	15.6 pg/mL
# 7	250 μL of 6	250 μL	7.8 pg/mL

Positive Control - Reconstitute the Positive Control with refer to lot specific of Dilution Buffer.

Detection Antibody Concentrate - Reconstitute the Detection Antibody Concentrate with refer to lot specific of **Antibody & HRP Diluent Solution** to produce a 10-fold concentrated stock solution. Pipette 9.45 mL of **Antibody & HRP Diluent Solution** into a 15 mL centrifuge tube and transfer 1.05 mL of 10-fold concentrated stock solution to prepare working solution.

Streptavidin-HRP Conjugate - Pipette 11.88 mL of **Antibody & HRP Diluent Solution** into a 15 mL centrifuge tube and transfer 120 μL of 100-fold

concentrated stock solution to prepare working solution (**protect from light**).

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Add 100 μ L per well of **Dilution Buffer** to blank wells.
3. Add 100 μ L of **Standard dilutions** in reverse order of serial dilution, **samples**, or **positive control** per well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
4. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with **1x Wash Buffer** (300 μ L) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
5. Add 100 μ L of **Detection Antibody working solution** to each well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
6. Repeat the aspiration/wash as in step 4.
7. Add 100 μ L of **Streptavidin-HRP Conjugate working solution** to each well. Incubate for 60 minutes on microplate shaker at room temperature. **Protect from light.**
8. Repeat the aspiration/wash as in step 4.
9. Add 100 μ L of **Substrate Solution** to each well. Incubate for refer to lot specific on microplate shaker at room temperature. **Protect from light.**
10. Add 100 μ L of **Stop Solution** to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
11. Determine the optical density of each well using a microplate reader set to 450 nm within 3 min.

CALCULATION OF RESULTS

Average the duplicate readings for each standard, positive control and sample, and subtract the average zero standard optical density. Create a standard curve by reducing the data using computer software capable of generating a log-log curve fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the LOX-1 concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

Readings of sample concentration exceeding that of the standard 500 pg/mL may result in inaccurate, low human sLOX-1 levels. Such samples require further external pre-dilution according to expected human sLOX-1 values with Dilution Buffer in order to precisely quantify the actual human sLOX-1 level.

SPECIFICITY

PROTEINS	CROSS-REACTIVITY
Human soluble LOX-1 (HEK293 derived)	100%
Human soluble LOX-1 (NS0 derived)	100%
Mouse soluble LOX-1	0
Human CD36	0
Human sRAGE	0
Human CD94	0
Human LDL	0
Human VLDL	0

The human LOX-1 extracellular domain recombinant derived from E. Coli or sf21 cells may not be detected by this ELISA Kit.

TYPICAL DATA

This standard curve is provided for demonstration only. A new standard curve should be generated for each set of samples assayed.

STANDARD (PG/ML)	AVERAGE OD450 (CORRECTED)*
Blank	0 (refer to lot)
7.8	0.043
15.6	0.091
31.25	0.179
62.5	0.288
125	0.584
250	1.119
500	1.999

SUMMARY OF ASSAY PROCEDURE

PREPARE REAGENTS, SAMPLES AND STANDARDS
↓
Add 100 µL of standard dilutions, samples, or positive control to each well. Incubate 2 hours on plate shaker at RT.
↓
Aspirate and wash 4 times.
↓
Add 100 µL Detection Antibody working solution to each well. Incubate 2 hours on plate shaker at RT.
↓
Aspirate and wash 4 times.
↓
Add 100 µL Streptavidin-HRP Conjugate working solution to each well. Incubate 60 min on plate shaker at RT. Protect from light.
↓
Aspirate and wash 4 times.
↓
Add 100 µL Substrate Solution to each well. Incubate refer to lot specific on plate shaker at RT. Protect from light.
↓
Add 100 µL Stop Solution to each well. Read at 450nm within 3 min.