

## HUMAN CTRP4 ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF  
HUMAN CTRP4 CONCENTRATIONS IN SERUM AND  
EDTA PLASMA



### PRODUCT INFORMATION:

ELISA NAME	HUMAN CTRP4 ELISA KIT
Catalog No.	SK00090-06
Lot No.	
Formulation	96 T
Standard range	1.56 - 100 ng/mL
Sensitivity	0.78 ng/mL
Sample Volume	100 µL
Dilution Factor	<b>Optimal dilutions should be determined by each laboratory for each application</b>
Sample Type	Serum, EDTA Plasma
Specificity	Human CTRP4
Calibration	Human CTRP4 Recombinant
Intra-assay Precision	6 - 8%
Inter-assay Precision	10 - 12%
Storage	2 - 8° C
This kit contains sufficient materials to run 35 samples duplicated provided that assay is run according to protocol.	

ALWAYS REFER TO LOT SPECIFIC PROTOCOL PROVIDED WITH EACH KIT FOR INSTRUCTIONS. PROTOCOL MUST BE READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

ORDER CONTACT:  
AVISCERA BIOSCIENCE, INC.  
2348 WALSH AVE., SUITE C  
SANTA CLARA, CA 95051  
USA

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: [Info@AvisceraBioscience.com](mailto:Info@AvisceraBioscience.com)

Website: [www.AvisceraBioscience.com](http://www.AvisceraBioscience.com)

## DESCRIPTION

This CTRP4 (H) ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human CTRP4 from serum and plasma in a sandwich ELISA format.

This immunoassay contains recombinant human CTRP4 and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural CTRP4 samples.

## ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with an antibody specific for human CTRP4. The capture antibody can bind to the human CTRP4 in the standard and samples. After washing the plate of any unbound substances, an antibody against human CTRP4 is added to the wells. After another washing of the plate, an Anti Rabbit IgG-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution is added to the wells and color develops in direct proportion to the amount of human CTRP4 bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

## PROCEDURAL LIMITATIONS

\_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

\_This ELISA kit should not be used beyond the expiration date on the kit label.

\_Do not mix reagents with those from other lots or sources.

\_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

\_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay.

\_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

\_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

## COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
<b>CTRP4 Microplate</b> - 96 well polystyrene microplate coated with an antibody against human CTRP4.	<b>090-06-01</b>	<b>1 plate</b>
<b>CTRP4 Standard</b> – 200 ng/vial of recombinant human soluble CTRP4 in a buffered protein base with preservative; lyophilized.	<b>090-06-02</b>	<b>1 vial</b>
<b>Detection Antibody Concentrate</b> – 1.05 mL/vial, 10-fold concentrate of purified antibody against human CTRP4 with preservative; lyophilized.	<b>090-06-03</b>	<b>1 vial</b>
<b>Positive Control</b> – one vial of recombinant human CTRP4; lyophilized.	<b>090-06-04</b>	<b>1 vial</b>
<b>Anti Rabbit IgG-HRP Conjugate</b> - 120 µL/vial, 100-fold concentrated solution of Anti Rabbit IgG conjugate to HRP.	<b>ARIGHRP</b>	<b>1 vial</b>
<b>Dilution Buffer</b> – 60 mL of buffered protein based solution with preservative.	<b>DB18B</b>	<b>1 bottle</b>
<b>HRP Diluent Solution</b> – 12 mL of buffered protein based solution with preservative.	<b>DB08</b>	<b>1 bottle</b>
<b>Wash Buffer</b> - 50 mL of 10-fold concentrated buffered surfactant, with preservative.	<b>WB01</b>	<b>1 bottle</b>
<b>TMB Substrate Solution</b> - 11 mL of substrate solution.	<b>TMB01</b>	<b>1 bottle</b>
<b>Stop Solution</b> - 11 mL of 0.5M HCl.	<b>S-STOP</b>	<b>1 bottle</b>
<b>Plate Sealer</b>	<b>EAPS</b>	<b>1</b>
<b>Plastic Pouch</b>	<b>P01</b>	<b>1</b>

## STORAGE

**Unopened Kit:** Store at 2 - 8° C for up to 6 months. For longer storage, unopened Standard, Positive Control and Detection Antibody Concentrate should be stored at -20° C or -70° C. Do not use kit past expiration date.

**Opened / Reconstituted Reagents:** Reconstituted Standard (stock) solution and Detection Antibody concentrated solution SHOULD BE STORED at -20° C or -70° C for up to one month. Anti Rabbit IgG-HRP Conjugate 100-fold concentrated solution (protect

from light) and other components may be stored at 2 - 8° C for up to 6 months. Do not freeze Substrate Solution and HRP.

**Microplate Wells:** Return unused wells to the plastic pouch with the desiccant pack. Microplate may be stored for up to 6 months at 2 - 8° C.

### ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

### PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

### SAMPLE COLLECTION AND STORAGE

**Serum** - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

**Plasma** - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

**Optional: Use Aprotinin (enzyme inhibitor) for ALL sample collection to prevent sample degradation. 0.5 TIU per ml of sample solution.**

### SAMPLE PREPARATION

Serum and plasma samples do not require dilution.

**Optimal dilutions should be determined by each laboratory for each application.**

**Use polypropylene test tubes.**


### REAGENT PREPARATION

**Bring all reagents to room temperature before use.**

**Wash Buffer** - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of 1x Wash Buffer.

**CTRP4 Standard** - Reconstitute the CTRP4 standard with 1 mL of Dilution Buffer. This reconstitution produces a stock solution of 200 ng/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 250 µL of Dilution Buffer into tubes #1-7. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The **100 ng/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 ng/mL).

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
stock	powder	1 ml	200 ng/ml
# 1	250µl of stock	250µl	100 ng/ml
# 2	250µl of 1	250µl	50 ng/ml
# 3	250µl of 2	250µl	25 ng/ml
# 4	250µl of 3	250µl	12.5 ng/ml
# 5	250µl of 4	250µl	6.25 ng/ml
# 6	250µl of 5	250µl	3.125 ng/ml
# 7	250µl of 6	250µl	1.56 ng/ml



**Positive Control** - Reconstitute the Positive Control with 1 mL of Dilution Buffer. **Note:** Positive Control could be reused within a few days if stored at -20° C or -70° C.

**Detection Antibody Concentrate** - Reconstitute the Detection Antibody Concentrate with 1.05 mL of Dilution Buffer to produce a 10-fold concentrated

stock solution. Pipette 9.45 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 1.05 mL of 10-fold concentrated stock solution to prepare working solution. **Note:** Detection Antibody Working Solution should be prepared 2 hours prior to use.

**Anti Rabbit IgG-HRP Conjugate** - Transfer 120 µL of 100-fold concentrated **Anti Rabbit IgG-HRP Conjugate** stock solution to 11.88 mL of **HRP Diluent Solution (DB08)** to prepare working solution. **Note:** *1x working solution of Anti Rabbit IgG-HRP Conjugate should be used within a few days (protect from light).*

## ELISA PROTOCOL

**Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.**

1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess microplate strips from the plate frame, return them to the plastic pouch with the desiccant pack.
3. Add 100 µL of **Dilution Buffer** to Blank wells.
4. Add 100 µL of **Standard dilutions** in reverse order of serial dilution, **sample**, or **positive control** per well. Cover with plate sealer. Incubate for 2 hours on microplate shaker (250-300rpm) at room temperature.
5. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with **1x Wash Buffer** (300 µL) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
6. Add 100 µL of **Detection Antibody working solution** to each well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
7. Repeat the aspiration/wash as in step 5.
8. Add 100 µL of **Anti Rabbit IgG-HRP Conjugate working solution** to each well. Incubate for 45 minutes on microplate shaker at room temperature. **Protect from light.**
9. Repeat the aspiration/wash as in step 5.

10. Add 100 µL of Substrate Solution to each well. Incubate for 5-10 minutes on microplate shaker at room temperature. **Protect from light.**
11. Add 100 µL of Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Determine the optical density of each well within 15 minutes, using a microplate reader set to 450 nm.

## CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

## TYPICAL DATA

This standard curve data is provided for demonstration only. A new standard curve should be generated for each set of samples assayed.









STANDARD (NG/ML)	AVERAGE OD450 (CORRECTED)
Blank	0 (0.093)
1.56	0.046
3.125	0.087
6.25	0.154
12.5	0.316
25	0.618
50	1.100
100	1.579

- Lot No.:
- Positive Control:

**SPECIFICITY**

PROTEINS	CROSS-REACTIVITY
Human CTRP4	100%
Human CTRP3	0
Human CTRP1	0
Human CTRP9	0
Human CTRP6	0
Human Adiponectin	0

**SUMMARY OF ASSAY PROCEDURE**

PREPARE REAGENTS, SAMPLES AND STANDARDS

Add 100 µL of standard dilutions, samples, or positive control to each well. Incubate 2 hours on the plate shaker at RT. Prepare Detection Antibody Working Solution.

Aspirate and wash 4 times.

Add 100 µL Detection Antibody working solution to each well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µL Anti Rabbit IgG HRP conjugate working solution to each well. Incubate 45 minutes on the plate shaker at RT. <b>Protect from light.</b>

Aspirate and wash 4 times.

Add 100 µL Substrate Solution to each well. Incubate 5-10 min on the plate shaker at RT. <b>Protect from light.</b>

Add 100 µL Stop Solution to each well. Read 450nm within 15 min.