

HUMAN SERUM PARAOXONASE 1 (PON1) ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF
HUMAN SERUM PARAOXONASE 1
CONCENTRATIONS IN SERUM AND PLASMA



**ALWAYS REFER TO LOT SPECIFIC PROTOCOL
PROVIDED WITH EACH KIT FOR
INSTRUCTIONS. PROTOCOL MUST BE
READ BEFORE USING THIS PRODUCT.**

**FOR RESEARCH USE ONLY. NOT FOR USE IN
DIAGNOSTIC PROCEDURES.**

PURCHASE INFORMATION:

ELISA NAME	HUMAN SERUM PARAOXONASE 1 (PON1) ELISA
Catalog No.	SK00141-01
Lot No.	
Formulation	96 T
Standard range	1.56 - 100 ng/mL
Sensitivity	100 pg/mL
Sample require	100 µL
Dilution Factor	Optimal dilutions should be determined by each laboratory for each application
Sample Type	Serum, EDTA Plasma
Specificity	Human Serum Paraoxonase 1
Calibration	Human Serum Paraoxonase 1 recombinant
Intra-assay Precision	4 - 6%
Inter-assay Precision	8 - 12%
Storage	2 – 8° C
This kit contains sufficient materials to run 35 samples duplicated provided that assay is run according to protocol.	

ORDER CONTACT:

**AVISCERA BIOSCIENCE, INC.
2348 Walsh Ave., Suite C
Santa Clara, CA 95051
USA**

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: Sales@AvisceraBioscience.com

www.AvisceraBioscience.com

DESCRIPTION

This Human Serum Paraoxonase 1 ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human Paraoxonase 1 from serum and plasma in a sandwich ELISA format.

This immunoassay contains recombinant human Paraoxonase 1 and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural Paraoxonase 1 samples.

ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with an antibody specific for human Paraoxonase 1. The capture antibody can bind to the human Paraoxonase 1 in the standard and samples. After washing the plate of any unbound substances, an antibody against human Paraoxonase 1 is added to the wells. After another washing of the plate, Anti Rabbit IgG-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution (TMB) is added to the wells and color develops in direct proportion to the amount of human Paraoxonase 1 bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
Paraoxonase 1 Microplate - 96 well microplate coated with an antibody specific for human Paraoxonase 1.	141-01-01	1 plate
Paraoxonase 1 Standard – 100 ng/vial of lyophilized recombinant human Paraoxonase 1.	141-01-02	1 vial
Detection Antibody – 1.05 mL/vial of 10-fold concentrate of lyophilized antibody against human Paraoxonase 1.	141-01-03	1 vial
Positive Control – one vial of lyophilized recombinant human Paraoxonase 1.	141-01-04	1 vial
Anti Rabbit IgG-HRP Conjugate - 120 µL/vial of 100-fold concentrated solution of Goat Anti Rabbit IgG conjugate to HRP.	ARIGHRP	1 vial
Dilution Buffer – 60 mL of buffered solution with preservative.	DB08	2 bottles
Wash Buffer – 50 mL of 10-fold concentrated buffered surfactant with preservative.	WB01	1 bottle
TMB Substrate Solution – 11 mL of TMB substrate solution.	TMB01	1 bottle
Stop Solution – 11 mL of 0.5M HCl.	S-STOP	1 bottle
Plate Sealer	EAPS	1
Plastic Pouch	P01	1

STORAGE

Unopened Kit: Store at 2 – 8° C for up to 8 months. For longer storage, unopened Standard, Positive Control and Detection Antibody Concentrate should be stored at -20° C or -70° C. Do not use kit past expiration date.

Opened / Reconstituted Reagents: Reconstituted Standard (stock) and Detection Antibody concentrated solution SHOULD BE STORED at -20° C or -70° C for up to one month. Anti Rabbit IgG-HRP Conjugate 100-fold concentrated solution (protect from light) and other components may be stored at 2 – 8° C for up to 8 months.

Microplate Wells: Return unused wells to the plastic pouch with the desiccant pack. Microplate may be stored for up to 6 months at 2 – 8° C after opening.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

SAMPLE COLLECTION AND STORAGE

Serum - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Plasma - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Optional: Use Aprotinin (enzyme inhibitor) for ALL sample collection to prevent sample degradation. 0.5 TIU per ml of sample solution.

SAMPLE PREPARATION

Serum samples may not require dilution. Plasma samples may require an 80-fold dilution. A suggested 10-fold dilution is 10 µL plasma sample + 90 µL Dilution Buffer. A suggested 80-fold dilution is 30 µL of 10-fold diluted plasma sample + 210 µL Dilution Buffer.

Optimal dilutions should be determined by each laboratory for each application with a pretest. Use polypropylene test tubes.


REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of 1x Wash Buffer.

Paraoxonase 1 Standard - Reconstitute the Paraoxonase 1 standard with 1.0 mL of Dilution Buffer. This reconstitution produces a stock solution of 100 ng/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 250 µL of Dilution Buffer into tubes #1 to #6. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The **100 ng/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 ng/mL).

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
Stock	powder	1.0 ml	100 ng/ml
# 1	250µl of stock	250µl	50 ng/ml
# 2	250µl of 1	250µl	25 ng/ml
# 3	250µl of 2	250µl	12.5 ng/ml
# 4	250µl of 3	250µl	6.25 ng/ml
# 5	250µl of 4	250µl	3.125 ng/ml
# 6	250µl of 5	250µl	1.563 ng/ml



Positive Control - Reconstitute the Positive Control with 0.5 mL of Dilution Buffer. **Note:** Positive Control solution could be reused within a few days if stored at -20° C or -70° C.

Detection Antibody - Reconstitute the Detection Antibody Concentrate with 1.05 mL of Dilution Buffer to produce a 10-fold concentrated stock solution. Pipette 9.45 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 1.05 mL of 10-fold concentrated stock solution to prepare working solution.

Anti Rabbit IgG-HRP Conjugate - Transfer 120 μL of 100-fold concentrated Anti Rabbit IgG-HRP Conjugate stock solution to 11.88 mL of Dilution Buffer to prepare working solution. **Note:** 1x working solution of Anti Rabbit IgG-HRP Conjugate should be used within a few days (**protect from light**).

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess microplate strips from the plate frame, return them to the plastic pouch with the desiccant pack.
3. Add 100 μL of Dilution Buffer to Blank wells.
4. Add 100 μL of Standard dilutions from #6 to #S, positive control and samples per well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
5. Aspirate and wash each well with 300 μL of **1x Wash Buffer** four times. After the last wash, aspirate any remaining 1x Wash Buffer, invert the plate and blot against clean paper towel(s).
6. Add 100 μL of Detection Antibody working solution to each well. Cover with the plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
7. Repeat the aspiration/wash as in step 5.
8. Add 100 μL of Anti Rabbit IgG-HRP Conjugate working solution to each well. Incubate for 60 minutes on microplate shaker at room temperature. **Protect from light.**
9. Repeat the aspiration/wash as in step 5.
10. Add 100 μL of Substrate Solution to each well. Incubate for 3-7 minutes on microplate shaker at room temperature. **Protect from light.**
11. Add 100 μL of Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Determine the optical density of each well within 15 minutes, using a microplate reader set to 450 nm.

CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

Calculation of samples with a concentration exceeding that of 100 ng/mL may result in inaccurate, low human Paraoxonase 1 levels. Such samples require further external predilution according to expected human Paraoxonase 1 values with Dilution Buffer in order to precisely quantify the actual human Paraoxonase 1 level.

SPECIFICITY

PROTEIN NAME	CROSS-REACTIVITY
Human PARAOXONASE 1	100%
Human MPO	0









TYPICAL STANDARD CURVE

This standard curve is for demonstration only. A new standard curve should be made for each set of samples assayed.

STANDARD (NG/ML)	AVERAGE OD450 (CORRECTED)
Blank	0 (0.073)
0.781 (optional)	0.039
1.563	0.089
3.125	0.152
6.25	0.250
12.5	0.521
25	0.819
50	1.118
100	1.516

- Lot No.:
- Positive control:

SUMMARY OF ASSAY PROCEDURE

PREPARE REAGENTS, SAMPLES AND STANDARDS

Add 100 µl of standard dilutions, samples, or positive control to each well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µl Detection Antibody working solution to each well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µl Anti Rabbit IgG-HRP conjugate working solution to each well. Incubate 60 min on the plate shaker at RT. Protect from light.

Aspirate and wash 4 times.

Add 100 µl Substrate Solution to each well. Incubate 3-7 min on the plate shaker at RT. Protect from light.

Add 100 µl Stop Solution to each well. Read 450nm within 15 min.