# HIGH SENSITIVITY CHEMERIN (HUMAN) ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF HUMAN CHEMERIN CONCENTRATIONS IN SERUM AND PLASMA



ALWAYS REFER TO LOT SPECIFIC PROTCOL PROVIDED WITH EACH KIT FOR INSTRUCTIONS. PROTOCOL MUST BE READ AND CHECK ALL ITEMS OF EACH KIT BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

## PRODUCT INFORMATION: THIS KIT IS FOR ONE TIME USE ONLY.

ELISA NAME	HIGH SENSITIVITY CHEMERIN (HUMAN) ELISA KIT KIT
Catalog No.	SK00171-02
Lot No.	
Formulation	96 T
Standard range	15.6-500 pg/mL
Sensitivity	7 pg/mL
Sample Volume	100 μL
Sample Type	Serum, Plasma
Dilution Factor	Optimal dilutions should be determined by each laboratory for each application
Specificity	Human chemerin
Specificity	
Calibration	Human chemerin recombinant
Calibration Intra-assay	recombinant

This kit contains sufficient materials to run approximately 35 samples duplicated provided that assay is run according to protocol.

### **ORDER CONTACT:**

AVISCERA BIOSCIENCE, INC. 2348 Walsh Ave., Suite C Santa Clara, CA 95051

**USA** 

Tel: (408) 982 0300 Fax: (408) 982 0301

Email: Sales@AvisceraBioscience.com

Info@AvisceraBioscience.com

www.AvisceraBioscience.net www.AvisceraBioscience.com

\_\_\_\_\_

### DESCRIPTION

This High Sensitivity Human Chemerin ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural Chemerin from serum samples and plasma in a sandwich ELISA format.

This immunoassay contains recombinant human Chemerin and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural human Chemerin samples.

### **ASSAY OVERVIEW**

This assay employs the quantitative sandwich enzyme immunoassay technique. The plate is precoated with a monoclonal antibody specific for human chemerin. The capture antibody can bind to the human chemerin in the standard and samples. After washing the plate of any unbound substances, a biotinylated antibody against human chemerin is added to the wells. After another washing of the plate, Streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution is added to the wells and color develops in direct proportion to the amount of human chemerin bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

### PROCEDURAL LIMITATIONS

\_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

\_This ELISA kit should not be used beyond the expiration date on the kit label.

\_Do not mix reagents with those from other lots or sources.

\_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

\_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

\_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

### COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
Chemerin Microplate - 96 well polystyrene microplate (12 strips of 8 wells) coated with a monoclonal antibody against human Chemerin.	401-01-01	1 plate
Human Chemerin Standard – refer to lot of recombinant human Chemerin in a buffered protein base with preservative; lyophilized.	171-02-02	1 vial
Detection Antibody Concentrate – 1.05 ml/vial, 10-fold concentrated of biotinylated antibody solution against human Chemerin with preservative	171-02-03	1 vial
Positive Control –one vial, of human Chemerin recombinant with preservative	171-02-04	1 vial
Streptavidin-HRP Conjugate - 120 µl/vial, 100- fold concentrated solution of Streptavidin conjugate to HRP.	SAHRP	1 vial
<b>Dilution Buffer</b> – 40 mL of buffered protein based solution with preservative.	DB01	2 bottles
HRP Diluent Solution – 12 mL of buffered protein based solution with preservative.	DB08B	1 bottle
Wash Buffer - 50 mL of 10- fold concentrated buffered surfactant, with preservative.	WB01	1 bottle
TMB Substrate Solution - 11 mL of TMB substrate solution.	TMB01	1 bottle
Stop Solution - 11 mL of 0.5M HCl solution.	S-STOP	1 bottle
Plate Sealer	EAPS	1 piece
Plastic Pouch	P01	1 piece

### **STORAGE**

**Unopened Kit:** Store at 2 – 8° C for up to 1 month. For longer storage up to 10 months, unopened Standard, Dilution Buffer and HRP Diluent Solution should be stored at -20° C. Streptavidin-HRP Conjugate and TMB Substrate Solution should be

\_\_\_\_\_

stored only at 2-8° C. Do not use kit past expiration date.

### ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

### **PRECAUTION**

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

### SAMPLE PREPARATION

Serum or plasma samples may require at least a 200-fold or 400-fold dilution. A 20-fold dilution is 10  $\mu L$  sample + 190  $\mu L$  of Dilution Buffer. Final to make a 200-fold dilution is 30  $\mu L$  of 20-fold diluted sample solution + 270  $\mu L$  of Dilution Buffer. Finally, to make a 400-fold dilution is 15  $\mu L$  of 20-fold diluted sample solution + 285  $\mu L$  of Dilution Buffer.

Optimal dilutions should be determined by each laboratory for each application with a sample pretest.

Use polypropylene test tubes.

### **REAGENT PREPARATION**

Bring all reagents to room temperature before use. Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of 1x Wash Buffer.

Chemerin Standard - Reconstitute the Chemerin standard with refer to lot of Dilution Buffer. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 450  $\mu L$  of Dilution Buffer into the tube #1. Pipette 250  $\mu L$  of Dilution Buffer into the tube #2 to

#6. Use the stock solution to produce a dilution

series (below). Mix each tube thoroughly before the next transfer. The **500 pg/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 pg/mL).

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
stock	powder	Refer to lot	2000 pg/ml
optional	125 μl of stock	125 μΙ	1000 pg/ml
#1	150µl of stock	450µl	500 pg/ml
# 2	250µl of 2	250μΙ	250 pg/ml
#3	250µl of 3	250μΙ	125 pg/ml
# 4	250µl of 4	250μΙ	62.5 pg/ml
# 5	250μl of 5	250μΙ	31.25 pg/ml
# 6	250μl of 6	250μΙ	15.6 pg/ml

**Positive Control** - Reconstitute the Positive Control with refer to lot of Dilution Buffer.

**Detection Antibody** - Reconstitute the Detection Antibody Concentrate with 1.05 mL of Dilution Buffer to prepare 10-fold concentrated stock. Pipette 9.45 mL of Dilution Buffer into a 15 ml centrifuge tube and transfer 1.05 mL of 10-fold concentrated stock solution to prepare working solution.

Streptavidin-HRP Conjugate - Pipette 11.88 mL of HRP Diluent solution (DB08B) into a 15 mL centrifuge tube and transfer 120  $\mu$ L of 100-fold concentrated stock solution to prepare working solution (protect from light).

### **ELISA PROTOCOL**

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

- 1. Prepare all reagents and working standards as directed in the previous sections.
- 2. Add 100 μL per well of Dilution Buffer to Blank wells.
- 3. Add 100  $\mu$ L of standard dilutions in reverse order of serial dilution, samples, or positive control per well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.

- 4. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with 1x Wash Buffer (300 μL) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
- 5. Add 100  $\mu$ L of Detection Antibody working solution to each well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
- 6. Repeat the aspiration/wash as in step 4.
- Add 100 μL of Streptavidin-HRP Conjugate working solution to each well. Incubate for 60 minutes on microplate shaker at room temperature. Protect from light.
- 8. Repeat the aspiration/wash as in step 4.
- 9. Add 100  $\mu$ L of Substrate Solution to each well. Incubate for refer to lot on microplate shaker at room temperature. **Protect from light.**
- 10. Add 100  $\mu$ L of Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
- 11. Determine the optical density of each well using a microplate reader set to 450 nm within 3 minutes.

### **CALCULATION OF RESULTS**

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit or a 4-parameter curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

### **SPECIFICITY**

PROTEINS	CROSS-REACTIVITY (%)
Human Chemerin	100
Mouse Chemerin	0
Human Visfatin	0
Human Resistin	0
Human Adiponectin	0

### TYPICAL STANDARD CURVE

This standard curve is provided for demonstration only. A new standard curve should be generated for each set of samples assayed.

STANDARD (PG/ML)	AVERAGE OD450 (CORRECTED)
Blank	0 (refer to lot)
15.6	0.032
31.25	0.068
62.5	0.155
125	0.387
250	0.945
500	2.252
1000 (optional)	3.341

### **SUMMARY OF ASSAY PROCEDURE**

# Add 100 µl of standard dilutions, samples, or positive control to the well. Incubate 2 hours on the plate shaker at RT. Aspirate and wash 4 times. Add 100 µl Detection Antibody working solution to each well. Incubate 2 hours on the plate shaker at RT. Aspirate and wash 4 times. Aspirate and wash 4 times. Aspirate and wash 4 times. Add 100 µl Streptavidin-HRP conjugate working solution to each well. Incubate 60 min on the plate shaker at RT. Protect from light. Aspirate and wash 4 times. Add 100 µl Substrate solution to each well. Incubate refer to lot on the plate shaker at RT. Protect from light.

Add 100  $\mu$ l Stop Solution to each well. Read at 450

nm within 3 minutes.