

URINARY SOLUBLE KIDNEY INJURY MOLECULE-1 (KIM-1) (HUMAN) ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF
HUMAN KIM-1 CONCENTRATIONS IN URINE,
SERUM AND PLASMA



ALWAYS REFER TO LOT SPECIFIC PROTOCOL
PROVIDED WITH EACH KIT FOR
INSTRUCTIONS. PROTOCOL MUST BE
READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN
DIAGNOSTIC PROCEDURES.

PRODUCT INFORMATION:

THIS KIT IS FOR ONE TIME USE ONLY.

ELISA NAME	HUMAN SOLUBLE KIM-1 ELISA KIT
Catalog No.	SK000186-08
Formulation	96 T
Lot No.	
Standard range	25 ~ 3200 pg/mL
Sensitivity	10 pg/mL
Sample Volume	100 µL
Dilution Factor	Optimal dilutions should be determined by each laboratory for each application
Sample Type	Urine, Serum, EDTA Plasma,
Specificity	Human
Calibration	Human KIM-1 Extracellular domain recombinant
Intra-assay Precision	4 - 6%
Inter-assay Precision	8 - 10%
Storage	2 – 8° C for 1 month. See page 2-3 for detail
This kit contains sufficient materials to run approximately 40 samples duplicated provided that assay is run according to protocol.	

Order Contact:

AVISCERA BIOSCIENCE, INC.

2348 Walsh Ave., Suite C

Santa Clara, CA 95051

USA

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: Sales@AvisceraBioscience.com

Info@AvisceraBioscience.com

www.AvisceraBioscience.com

DESCRIPTION

This Human Soluble KIM-1 ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human Soluble KIM-1 from urine, serum and plasma in a sandwich ELISA format.

This immunoassay contains recombinant human KIM-1 and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural KIM-1 samples.

ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with an antibody specific for human soluble KIM-1. The capture antibody can bind to the human soluble KIM-1 in the standard and samples. After washing the plate of any unbound substances, the biotinylated antibody against human KIM-1 is added to the wells. After another washing of the plate, Streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution (TMB) is added to the wells and color develops in direct proportion to the amount of human KIM-1 bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Each laboratory must determine the optimal dilution factors for the samples being assayed.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
KIM-1 Microplate – 96 well microplate coated with a monoclonal antibody specific for human KIM-1.	186-08-01	1 plate
KIM-1 Standard – refer to lot of lyophilized recombinant human KIM-1.	186-08-02	1 vial
Detection Antibody Concentrate – refer to lot of 10-fold concentrate of lyophilized antibody against human KIM-1.	186-08-03	1 vial
Positive Control – one vial of lyophilized recombinant human KIM-1.	158-08-04	1 vial
Streptavidin-HRP Conjugate – 120 µL/vial of 100-fold concentrated solution of Streptavidin-HRP conjugate.	SAHRP	1 vial
Dilution Buffer – 45 mL of buffered solution with preservative.	DB01	1 bottle
HRP Diluent Solution – 12 mL of buffered solution with preservative.	DB08C	1 bottle
Wash Buffer – 50 mL of 10-fold concentrated buffered surfactant with preservative.	WB01	1 bottle
TMB Substrate Solution – 11 mL of TMB substrate solution.	TMB01	1 bottle
Stop Solution – 11 mL of 0.5M HCl.	S-STOP	1 bottle
Plate Sealer	EAPS	1 piece
Plastic Pouch	P01	1 piece

STORAGE

Unopened Kit: Store at 2 – 8° C for up to 1month. For longer storage for up to 10 months, unopened Standard, Positive Control, Detection Antibody Concentrate, Dilution Buffer and Antibody & HRP Diluent Solution should be stored at -20° C. Streptavidin-HRP Conjugate and TMB Substrate Solution should be stored only at 2 ~ 8°C. Do not use kit past expiration date.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (350 – 400 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

SAMPLE COLLECTION AND STORAGE

Serum – Use a serum separator tube (SST). Allow blood to clot for 30 minutes. Centrifuge at 1000 x g for 15 minutes and collect serum. Assay samples immediately or aliquot and store at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Plasma – Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge at 1000 x g for 15 minutes and collect plasma. Assay samples immediately or aliquot and store at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Optional: Use Aprotinin (enzyme inhibitor) for ALL sample collection to prevent sample degradation. 0.5 TIU per mL of sample solution.

SAMPLE PREPARATION

Optimal dilutions should be determined by each laboratory for each application with a pretest. Use polypropylene test tubes.

REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer – If crystals have formed in the concentrate, warm bottle in a water bath until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into 450 mL distilled or deionized water to make 500 mL of 1x Wash Buffer.

KIM-1 Standard – Reconstitute the KIM-1 standard with refer to lot of Dilution Buffer. Allow the stock standard to sit for at least 15 minutes with gentle agitation until completely dissolved prior to making standard dilutions (see below). Mix each tube thoroughly before the next transfer. Create a standard curve using a following serial dilution in Dilution Buffer with a high standard of **3200 pg/mL** is recommended.

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
Stock	powder	Refer to lot	xxx
# 1	Refer to lot	Refer to lot	3200 pg/mL
# 2	150 μL of 1	450 μL	800 pg/mL
# 3	150 μL of 2	450 μL	200 pg/mL
# 4	150 μL of 3	450 μL	50 pg/mL
# 5	125 μL of 4	125 μL	25 pg/mL

Positive Control - Reconstitute the Positive Control with refer to lot Dilution Buffer.

Detection Antibody - Reconstitute the Detection Antibody Concentrate with refer to lot of Dilution Buffer to produce a 10-fold concentrated stock solution. Allow the concentrated solution to sit for at least 5 minutes until completely dissolved. Pipette 9.45 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 1.05 mL of 10-fold concentrated stock solution to prepare working solution.

Streptavidin HRP Conjugate - Pipette 9.395 mL of **HRP Diluent Solution (DB08C)** into a 15 mL centrifuge tube and transfer 105 μL of 100-fold concentrated stock solution to prepare working solution (**protect from light**). **DO NOT FREEZE.**

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents, standard dilutions, positive control and samples as directed previously.
2. Add 100 μL per well of **Dilution Buffer** to Blank wells.

3. Add 100 μ L per well of **Standard dilutions, samples, or positive control**. Cover with plate sealer and incubate for 2 hours on microplate shaker at room temperature.
4. Aspirate and wash each well with 300 μ L of **1x Wash Buffer** four times. After the last wash, aspirate any remaining 1x Wash Buffer, invert the plate and blot against clean paper towel(s).
5. Add 100 μ L per well of **Detection Antibody working solution**. Cover with plate sealer and incubate for 2 hours on microplate shaker at room temperature.
6. Repeat the aspiration and wash as in step 4.
7. Add 100 μ L per well of **Streptavidin HRP Conjugate working solution**. Cover with plate sealer and incubate for 60 minutes on microplate shaker at room temperature. **Protect from light.**
8. Repeat the aspiration and wash as in step 4.
9. Add 100 μ L per well of **Substrate Solution**. Incubate for refer to lot on microplate shaker at room temperature. **Protect from light.**
10. Add 100 μ L per well of **Stop Solution**. The color in the wells should change from blue to yellow. If the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
11. Read plate using a microplate reader set to 450 nm within 3 minutes.

CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

SPECIFICITY

Protein	Cross-reactivity (%)
Human KIM-1	100
Human KIM-4	0
Mouse KIM-1	0

TYPICAL STANDARD CURVE

This standard curve is for demonstration only. A new standard curve should be made for each set of samples assayed.

STANDARD (PG/ML)	AVERAGE OD450 NM (CORRECTED)
Blank	0 (refer to lot)
25	0.045
50	0.099
200	0.285
800	1.058
3200	2.479
12800 (optional)	3.139

SUMMARY OF ASSAY PROCEDURE

PREPARE REAGENTS, SAMPLES AND STANDARD DILUTIONS
↓
Add 100 μ L of standard dilutions, samples or positive control. Cover with plate sealer and incubate 2 hours on microplate shaker at RT.
↓
Aspirate and wash 4 times.
↓
Add 100 μ L per well of Detection Antibody working solution. Cover with plate sealer and incubate 2 hours on microplate shaker at RT.
↓
Aspirate and wash 4 times.
↓
Add 100 μ L per well of Streptavidin HRP Conjugate working solution. Cover with plate sealer and incubate 60 minutes on microplate shaker at RT. Protect from light.
↓
Aspirate and wash 4 times.
↓
Add 100 μ L per well of Substrate Solution. Incubate refer to lot on microplate shaker at RT. Protect from light.
↓
Add 100 μ L per well of Stop Solution. Read at 450 nm within 3 min.

The research samples were diluted by DB01 and assayed by
Urine KIM-1 (Human) ELISA Kit SK00186-08

Sample Type	Dilution Factor	Assayed (pg/mL)	Final (pg/mL)	Recovery (%)
Human Urine	1 X	961.446	961.446	100
Human Urine	2 X	459.108	918.216	95.5
Human Urine	4 X	236.482	945.928	98.4