

MOUSE SOLUBLE CD36 ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF MOUSE CD36 CONCENTRATIONS IN CELL CULTURE SUPERNATES AND PLASMA



ALWAYS REFER TO LOT SPECIFIC PROTOCOL PROVIDED WITH EACH KIT FOR INSTRUCTIONS. PROTOCOL MUST BE READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

PRODUCT INFORMATION:

ELISA NAME	MOUSE sCD36 ELISA
Catalog No.	SK00196-03
Lot No.	
Formulation	96 T
Standard range	0.22 - 14 ng/mL
Sensitivity	0.11 ng/mL
Sample Volume	100 μ L
Dilution Factor	<i>Optimal dilutions should be determined by each laboratory for each application.</i>
Sample Type	EDTA Plasma, Cell Culture Supernates
Specificity	Mouse sCD36
Calibration	Mouse sCD36 recombinant
Intra-assay Precision	6 - 8%
Inter-assay Precision	10 - 12%
Storage	2-8° C
This kit contains sufficient materials to run 35 samples duplicated provided that assay is run according to protocol.	

ORDER CONTACT:

AVISCERA BIOSCIENCE, INC.
2348 WALSH AVE., SUITE C
SANTA CLARA, CA 95051
USA

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: info@AvisceraBioscience.com

Website: www.AvisceraBioscience.com

DESCRIPTION

This Mouse sCD36 ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural mouse sCD36 from cell culture supernates and plasma in a sandwich ELISA format.

This immunoassay contains recombinant mouse sCD36 and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural sCD36 samples.

ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with an antibody specific for mouse sCD36. The capture antibody can bind to the mouse sCD36 in the standard and samples. After washing the plate of any unbound substances, a biotinylated antibody against mouse sCD36 is added to the wells. After another washing of the plate, Streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution is added to the wells and color develops in direct proportion to the amount of mouse sCD36 bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
Mouse sCD36 Microplate - 96 well polystyrene microplate coated with a polyclonal antibody IgG against sCD36.	196-03-01	1 plate
sCD36 Standard – 14 ng/vial of recombinant mouse sCD36 in a buffered protein base with preservative; lyophilized.	196-03-02	1 vial
Detection Antibody Concentrate – 1.05 mL/vial, 10-fold concentrate of biotinylated polyclonal IgG against sCD36 with preservative; lyophilized.	196-03-03	1 vial
Positive Control - one vial of recombinant mouse sCD36 in a buffered protein base with preservative; lyophilized.	196-03-04	1 vial
Streptavidin-HRP Conjugate -60 µl/vial, 200-fold concentrated solution of Streptavidin conjugate to HRP.	SAHRP	1 vial
Dilution Buffer - 60 mL of buffered protein based solution with preservative.	DB06	1 bottle
Wash Buffer - 50 mL of 10-fold concentrated buffered surfactant, with preservative.	WB01	1 bottle
TMB Substrate Solution - 11 mL of substrate solution.	TMB01	1 bottle
Stop Solution – 11 mL of 0.5M HCl.	S-STOP	1 bottle
Plate Sealer	EAPS	1 piece
Plastic Pouch	P01	1 piece

STORAGE

Unopened Kit: Store at 2 - 8° C for up to 6 months. For longer storage, unopened Standard, Positive Control and Detection Antibody Concentrate should be stored at -20° C or -70° C. Do not use kit past expiration date.

Opened / Reconstituted Reagents: Reconstituted Standard (14 ng/mL) and Detection Antibody concentrated solution SHOULD BE STORED at -20° C or -70° C for up to one month. Streptavidin-HRP Conjugate 200-fold concentrated solution and other components may be stored at 2 – 8° C for up to 6 months.

Microplate Wells: Return unused wells to the plastic pouch with the desiccant pack. Microplate may be stored for up to 6 months at 2 - 8° C.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

SAMPLE COLLECTION AND STORAGE

Cell Culture Supernates - Remove particulates by centrifugation and assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Plasma - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Serum: CD36 was expressed in plates. Activation of plates may increase sCD36 release. Serum samples cannot be used for sCD36 assay.

Optional: Use Aprotinin (enzyme inhibitor) for ALL sample collection to prevent sample degradation. 0.5 TIU per ml of sample solution.


REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of 1x Wash Buffer.

sCD36 Standard - Reconstitute the sCD36 standard with 1 mL of Dilution Buffer. This reconstitution produces a stock solution of 14 ng/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 250 µL of Dilution Buffer into tubes #1 to #6. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The **14 ng/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 ng/mL).

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
stock	powder	1 ml	14 ng/ml
# 1	250µl of stock	250µl	7 ng/ml
# 2	250µl of 1	250µl	3.5 ng/ml
# 3	250µl of 2	250µl	1.75 ng/ml
# 4	250µl of 3	250µl	0.875 ng/ml
# 5	250µl of 4	250µl	0.437 ng/ml
# 6	250µl of 5	250µl	0.218 ng/ml



Positive Control- Reconstitute the Positive Control with 1.0 mL of Dilution Buffer. *Note: Positive Control could be reused within a few days if stored at -20° C to -70° C.*

Detection Antibody - Reconstitute the Detection Antibody Concentrate with 1.05 mL of Dilution Buffer to produce a 10-fold concentrated stock solution. Pipette 9.45 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 1.05 mL of 10-fold concentrated stock solution to prepare working solution.

Streptavidin-HRP Conjugate - Pipette 11.94 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 60 µl of 200-fold concentrated stock solution to prepare working solution. *Note: 1x working solution of Streptavidin-HRP Conjugate should be used within a few days.*

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess microplate strips from the plate frame, return them to the plastic pouch with the desiccant pack.
3. Add 100 µL of Dilution Buffer to Blank wells.
4. Add 100 µL of Standard dilutions, sample, or positive control per well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
5. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with 1x Wash Buffer (300 µL) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
6. Add 100 µL of Detection Antibody working solution to each well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
7. Repeat the aspiration/wash as in step 5.
8. Add 100 µL of Streptavidin-HRP Conjugate working solution to each well. Incubate for 40 minutes on microplate shaker at room temperature. **Protect from light.**
9. Repeat the aspiration/wash as in step 5.
10. Add 100 µL of Substrate Solution to each well. Incubate for 15-30 minutes at on microplate shaker at room temperature. **Protect from light.**
11. Add 100 µL of Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.

12. Determine the optical density of each well within 15 minutes, using a micro-plate reader set to 450 nm.

CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

TYPICAL DATA

This standard curve is provided for demonstration only. A new standard curve should be generated for each set of samples assayed.

STANDARD (PG/ML)	CORRECTED (450NM)
Blank	0 (0.013)
213.75	0.042
437.5	0.075
875	0.117
1750	0.256
3500	0.483
7000	0.864
14000	1.517

- Lot No.:
- Positive Control:









SPECIFICITY

PROTEIN NAME	CROSS-REACTIVITY%
Mouse CD36/Fc chimera	100
Human CD36/Fc chimera	15.6
Human CD320	0
Rat sRAGE	0

REFERENCES

1. Masson CJ, et al. Fatty acid- and cholesterol transporter protein expression along the human intestinal tract. PLoS One. 2010 Apr 29;5(4):e10380.
2. Marecki JC, et al. Hyperinsulinemia and ectopic fat deposition can develop in the face of hyperadiponectinemia in young obese rats. J Nutr Biochem. 2010 Apr 30. [Epub ahead of print]
3. Bell JA, et al. Lipid partitioning, incomplete fatty acid oxidation, and insulin signal transduction in primary human muscle cells: effects of severe obesity, fatty acid incubation, and fatty acid translocase/CD36 overexpression. J Clin Endocrinol Metab. 2010 Jul;95(7):3400-10. Epub 2010 Apr 28.
4. Sandoval JC, et al. Fenofibrate reduces postprandial hypertriglyceridemia in CD36 knockout mice. J Atheroscler Thromb. 2010 Jun 30;17(6):610-8. Epub 2010 Mar 30.
5. Abe T, et al. Key role of CD36 in Toll-like receptor 2 signaling in cerebral ischemia. Stroke. 2010 May;41(5):898-904. Epub 2010 Apr 1.
6. Steinbusch LK, et al. Differential regulation of cardiac glucose and fatty acid uptake by endosomal pH and actin filaments. Am J Physiol Cell Physiol. 2010 Jun;298(6):C1549-59. Epub 2010 Apr 7.

SUMMARY OF ASSAY PROCEDURE

Prepare reagents, samples and standards

Add 100 µl of standard dilutions, samples, or positive control to each well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µl Detection Antibody working solution to each well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µl Streptavidin-HRP conjugate working solution to each well. Incubate 40 min on the plate shaker at RT. Protect from light.

Aspirate and wash 4 times.

Add 100 µl Substrate Solution to each well. Incubate 15-30 min on the plate shaker at RT. Protect from light.

Add 100 µl Stop Solution to each well. Read 450nm within 15 min.