

HUMAN NOV/CCN3 ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF HUMAN NOV/CCN3 CONCENTRATIONS IN SERUM, PLASMA AND CELL CULTURE OR TISSUES



ALWAYS REFER TO LOT SPECIFIC PROTOCOL PROVIDED WITH EACH KIT FOR INSTRUCTIONS. PROTOCOL MUST BE READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

PRODUCT INFORMATION:

| ELISA NAME | HUMAN NOV/CCN3 ELISA |
|---|---|
| Catalog No. | SK00265-01 |
| Lot No. | |
| Formulation | 96 T |
| Standard range | 31.25 – 2000 pg/mL |
| Sensitivity | 15 pg/mL |
| Sample Volume | 100 µl |
| Sample Type | Serum, Plasma, Cell Culture or Tissue |
| Specificity | Human NOV/CCN3 |
| Calibration | Human NOV/CCN3 recombinant |
| Dilution Factor | 10 (Optimal dilutions should be determined by each laboratory for each application) |
| Intra-assay Precision | 4 - 6% |
| Inter-assay Precision | 8 - 10% |
| Storage | 2 – 8° C |
| This kit contains sufficient materials to run 35 samples duplicated provided that assay is run according to protocol. | |

Order Contact:

AVISCERA BIOSCIENCE, INC.
2348 Walsh Ave., Suite C
Santa Clara, CA 95051
USA

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: Sales@AvisceraBioscience.com

Info@AvisceraBioscience.com

www.AvisceraBioscience.com

DESCRIPTION

This Human NOV/CCN3 ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human NOV/CCN3 from cell culture or tissues, serum and plasma in a sandwich ELISA format.

This immunoassay contains recombinant human NOV/CCN3 and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural NOV/CCN3 samples.

ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with an antibody specific for human NOV/CCN3. The capture antibody can bind to the human NOV/CCN3 in the standard and samples. After washing the plate of any unbound substances, a biotinylated antibody against human NOV/CCN3 is added to the wells. After another washing of the plate, Streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution (TMB) is added to the wells and color develops in direct proportion to the amount of human NOV/CCN3 bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

| DESCRIPTION | CODE | QUANTITY |
|--|------------------|-----------------|
| NOV/CCN3 Microplate - 96 well microplate coated with an antibody specific for human NOV/CCN3. | 265-01-01 | 1 plate |
| NOV/CCN3 Standard – 2000 pg/vial of lyophilized recombinant human NOV/CCN3. | 265-01-02 | 1 vial |
| Detection Antibody Concentrate – 1.05 mL/vial of 10-fold concentrate of lyophilized biotinylated antibody against human NOV/CCN3. | 265-01-03 | 1 vial |
| Positive Control – one vial of lyophilized recombinant human NOV/CCN3. | 265-01-04 | 1 vial |
| Streptavidin-HRP Conjugate – 60 µL/vial of 200-fold concentrated solution of Streptavidin-HRP conjugate. | SAHRP | 1 vial |
| Dilution Buffer – 60 mL of buffered solution with preservative. | DB01 | 1 bottle |
| HRP Diluent Solution – 12 mL of buffered solution with preservative. | DB08 | 1 bottle |
| Wash Buffer – 50 mL of 10-fold concentrated buffered surfactant with preservative. | WB01 | 1 bottle |
| TMB Substrate Solution – 11 mL of TMB substrate solution. | TMB01 | 1 bottle |
| Stop Solution – 11 mL of 0.5M HCl. | S-STOP | 1 bottle |
| Plate Sealer | EAPS | 1 piece |
| Plastic Pouch | P01 | 1 piece |

STORAGE

Unopened Kit: Store at 2 – 8° C for up to 6 months. For longer storage, unopened Standard, Positive Control and Detection Antibody Concentrate should be stored at -20° C or -70° C. Do not use kit past expiration date.

Opened / Reconstituted Reagents: Reconstituted Standard (stock) solution and Detection Antibody concentrated solution SHOULD BE STORED at -20° C or -70° C for up to one month. Streptavidin-HRP Conjugate 200-fold concentrated solution and TMB

Substrate Solution can be stored at 2 – 8° C for up to 6 months (**DO NOT FREEZE** and **PROTECT FROM LIGHT**). All other components may be stored at 2 – 8° C for up to 6 months.

Microplate Wells: Return unused wells to the plastic pouch with the desiccant pack. Microplate may be stored for up to 6 months at 2 – 8° C.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

SAMPLE COLLECTION AND STORAGE

Cell Culture Supernates – Centrifuge and assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Serum – Use a serum separator tube (SST). Allow blood to clot for 30 minutes. Centrifuge at 1000 x g for 15 minutes and collect serum. Assay samples immediately or aliquot and store at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Plasma – Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge at 1000 x g for 15 minutes and collect plasma. Assay samples immediately or aliquot and store at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Optional: Use Aprotinin (enzyme inhibitor) for ALL sample collection to prevent sample degradation. 0.5 TIU per mL of sample solution.

SAMPLE PREPARATION

Serum and Plasma samples may require a 10-fold dilution. A suggested 10-fold dilution is 25 µL of sample + 225 µL of Dilution Buffer. This is only a

suggestion. A pretest is needed to determine the optimal dilutions for your samples.

Optimal dilutions should be determined by each laboratory for each application with a pretest. Use polypropylene test tubes.


REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of 1x Wash Buffer.

NOV/CCN3 Standard - Reconstitute the NOV/CCN3 standard with 1 mL of Dilution Buffer. This reconstitution produces a stock solution of 2000 pg/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 250 µL of Dilution Buffer into tubes #1 to #6. Use the stock solution to produce a dilution series (see below). Mix each tube thoroughly before the next transfer. The **2000 pg/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 pg/mL).

| TUBE | STANDARD | DILUTION BUFFER | CONCENTRATION |
|-------|-----------------|-----------------|---------------|
| stock | Powder | 1000 µl | 2000 pg/ml |
| # 1 | 250 µl of stock | 250 µl | 1000 pg/ml |
| # 2 | 250 µl of 1 | 250 µl | 500 pg/ml |
| # 3 | 250 µl of 2 | 250 µl | 250 pg/ml |
| # 4 | 250 µl of 3 | 250 µl | 125 pg/ml |
| # 5 | 250 µl of 4 | 250 µl | 62.5 pg/ml |
| # 6 | 250 µl of 5 | 250 µl | 31.25 pg/ml |



Positive Control – Reconstitute the Positive Control with 1 mL of Dilution Buffer. **Note:** Positive Control could be reused within a few days if stored at -20° C or -70° C.

Detection Antibody Concentrate - Reconstitute the Detection Antibody Concentrate with 1.05 mL of Dilution Buffer to produce a 10-fold concentrated stock solution. Pipette 9.45 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 1.05 mL of 10-fold concentrated stock solution to prepare working solution.

Streptavidin-HRP Conjugate - Pipette 11.94 mL of **HRP Diluent Solution (DB08)** into a 15 mL centrifuge tube and transfer 60 µL of 200-fold concentrated stock solution to prepare working solution. **Note:** 1x working solution of Streptavidin-HRP should be used within a few days (**protect from light**). **DO NOT FREEZE.**

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess microplate strips from the plate frame, return them to the plastic pouch with the desiccant pack.
3. Add 100 µL of Dilution Buffer to Blank wells.
4. Add 100 µL of **Standard dilutions, sample, or positive control** per well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
5. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with **1x Wash Buffer** (300 µL) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
6. Add 100 µL of **Detection Antibody working solution** to each well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.

7. Repeat the aspiration/wash as in step 5.
8. Add 100 µL of **Streptavidin-HRP Conjugate working solution** to each well. Incubate for 60 minutes on microplate shaker at room temperature. **Protect from light.**
9. Repeat the aspiration/wash as in step 5.
10. Add 100 µL of **Substrate Solution** to each well. Incubate for 1-10 minutes on a microplate shaker at room temperature. **Protect from light.**
11. Add 100 µL of **Stop Solution** to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Determine the optical density of each well within 15 minutes, using a microplate reader set to 450 nm.

CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

SPECIFICITY

| ADIOKINES | CROSS-REACTIVITY (%) |
|--------------------------|----------------------|
| HUMAN NOV/CCN3 | 100 |
| MOUSE NOV | 7.7% |
| HUMAN NOTCH-1/Fc CHIMERA | 0 |
| HUMAN S100-A4 | 0 |
| HUMAN WISP-1 | 0 |

TYPICAL STANDARD CURVE

This standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.

| NOV/CCN3 (PG/ML) | AVERAGE OD450 (CORRECTED)* |
|-----------------------------|---------------------------------------|
| Blank | 0 (0.118) |
| 31.25 | 0.031 |
| 62.5 | 0.059 |
| 125 | 0.101 |
| 250 | 0.183 |
| 500 | 0.352 |
| 1000 | 0.589 |
| 2000 | 1.102 |

- Lot No.:
- Positive Control:

SUMMARY OF ASSAY PROCEDURE**PREPARE REAGENTS, SAMPLES AND STANDARDS**

↓
Add 100 µL of standard dilutions, samples, or positive control to each well. Incubate 2 hours on the plate shaker at RT.

↓
Aspirate and wash 4 times.

↓
Add 100 µL Detection Antibody working solution to each well. Incubate 2 hours on the plate shaker at RT.

↓
Aspirate and wash 4 times.

↓
Add 100 µL Streptavidin-HRP conjugate working solution to each well. Incubate 60 minutes on the plate shaker at RT. **Protect from light.**

↓
Aspirate and wash 4 times.

↓
Add 100 µL Substrate Solution to each well. Incubate 1-10 minutes on the plate shaker at RT. **Protect from light.**

↓
Add 100 µL Stop Solution to each well. Read at 450nm within 15 minutes.