

## HUMAN EPIGEN (EPGN) ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF  
HUMAN EPGN CONCENTRATIONS IN SERUM  
AND EDTA PLASMA



ALWAYS REFER TO LOT SPECIFIC PROTOCOL  
PROVIDED WITH EACH KIT FOR  
INSTRUCTIONS. PROTOCOL MUST BE  
READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN  
DIAGNOSTIC PROCEDURES.

### PRODUCT INFORMATION:

**THIS KIT IS FOR ONE TIME USE ONLY.**

ELISA NAME	HUMAN EPIGEN ELISA KIT
Catalog No.	SK000340-06
Formulation	96 T
Lot No.	
Standard range	0.78 -50 ng/mL
Sensitivity	150 pg/mL
Sample Volume	100 µL
Dilution Factor	Optimal dilutions should be determined by each laboratory for each application
Sample Type	Serum, EDTA Plasma
Specificity	Human
Calibration	Human Epigen Rec.
Intra-assay Precision	4 - 6%
Inter-assay Precision	8 - 10%
Storage	2 – 8° C for 1 month. See page 2-3 for detail
This kit contains sufficient materials to run approximately 40 samples duplicated provided that assay is run according to protocol.	

### Order Contact:

AVISCERA BIOSCIENCE, INC.  
2348 Walsh Ave., Suite C  
Santa Clara, CA 95051  
USA

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: [Sales@AvisceraBioscience.com](mailto:Sales@AvisceraBioscience.com)

[Info@AvisceraBioscience.com](mailto:Info@AvisceraBioscience.com)

[www.AvisceraBioscience.com](http://www.AvisceraBioscience.com)

## DESCRIPTION

This Human Soluble Epigen (EPGN) ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human EPGN from serum and plasma in a sandwich ELISA format.

This immunoassay contains recombinant human Epigen and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural Epigen samples.

## ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with a monoclonal antibody specific for human Epigen. The capture antibody can bind to the human Epigen in the standard and samples. After washing the plate of any unbound substances, the biotinylated monoclonal antibody against human Epigen is added to the wells. After another washing of the plate, Streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution (TMB) is added to the wells and color develops in direct proportion to the amount of human Epigen bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

## PROCEDURAL LIMITATIONS

\_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

\_This ELISA kit should not be used beyond the expiration date on the kit label.

\_Do not mix reagents with those from other lots or sources.

\_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

\_Each laboratory must determine the optimal dilution factors for the samples being assayed.

\_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

\_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

## COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
<b>Epigen Microplate</b> – 96 well microplate coated with a monoclonal antibody specific for human Epigen.	<b>340-06-01</b>	<b>1 plate</b>
<b>EPIGEN Standard</b> – refer to lot of lyophilized recombinant human Epigen.	<b>340-06-02</b>	<b>1 vial</b>
<b>Detection Antibody Concentrate</b> – refer to lot of 10-fold concentrate of lyophilized antibody against human Epigen.	<b>340-06-03</b>	<b>1 vial</b>
<b>Positive Control</b> – one vial of lyophilized recombinant human Epigen. (Optional)	<b>340-06-04</b>	<b>1 vial</b>
<b>Streptavidin-HRP Conjugate</b> – 120 µL/vial of 100-fold concentrated solution of Streptavidin-HRP conjugate.	<b>SAHRP</b>	<b>1 vial</b>
<b>Dilution Buffer</b> – 45 mL of buffered solution with preservative.	<b>DB16</b>	<b>1 bottle</b>
<b>Antibody Diluent Solution</b> – 12 mL of buffered solution with preservative.	<b>DB01B</b>	<b>1 bottle</b>
<b>HRP Diluent Solution</b> – 12 mL of buffered solution with preservative.	<b>DB08C</b>	<b>1 bottle</b>
<b>Wash Buffer</b> – 50 mL of 10-fold concentrated buffered surfactant with preservative.	<b>WB02</b>	<b>1 bottle</b>
<b>TMB Substrate Solution</b> – 11 mL of TMB substrate solution.	<b>TMB01</b>	<b>1 bottle</b>
<b>Stop Solution</b> – 11 mL of 0.5M HCl.	<b>S-STOP</b>	<b>1 bottle</b>
<b>Plate Sealer</b>	<b>EAPS</b>	<b>1 piece</b>
<b>Plastic Pouch</b>	<b>P01</b>	<b>1 piece</b>

## STORAGE

**Unopened Kit:** Store at 2 – 8° C for up to 1month. For longer storage for up to 10 months, unopened Standard, Positive Control, Detection Antibody Concentrate, Dilution Buffer , Antibody & HRP Diluent Solution should be stored at -20° C.

Streptavidin-HRP Conjugate and TMB Substrate Solution should be stored only at 2 ~ 8°C. Do not use kit past expiration date.

### ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

### PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

### SAMPLE COLLECTION AND STORAGE

**Serum** – Use a serum separator tube (SST). Allow blood to clot for 30 minutes. Centrifuge at 1000 x g for 15 minutes and collect serum. Assay samples immediately or aliquot and store at ≤ -20° C. Avoid repeated freeze-thaw cycles.

**Plasma** – Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge at 1000 x g for 15 minutes and collect plasma. Assay samples immediately or aliquot and store at ≤ -20° C. Avoid repeated freeze-thaw cycles.

**Optional: Use Aprotinin (enzyme inhibitor) for ALL sample collection to prevent sample degradation. 0.5 TIU per mL of sample solution.**

### SAMPLE PREPARATION

Serum and plasma samples need to be diluted by 4-16 fold. **Optimal dilutions should be determined by each laboratory for each application with a pretest. Use polypropylene test tubes.**

### REAGENT PREPARATION

**Bring all reagents to room temperature before use.**

**Wash Buffer** – If crystals have formed in the concentrate, warm bottle in a water bath until the

crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into 450 mL distilled or deionized water to make 500 mL of 1x Wash Buffer.

**Dilution Buffer (DB16) - Dilution Buffer (DB16) is highly viscous, warm in 30 - 37° C water bath until liquid flows more freely.**

**Epigen Standard** – Reconstitute the Epigen standard with refer to lot of Dilution Buffer. Allow the stock standard to sit for at least 15 minutes with gentle agitation until completely dissolved prior to making standard dilutions (see below). Mix each tube thoroughly before the next transfer. Create a standard curve using a 2-fold serial dilution in Dilution Buffer with a high standard of **50 ng/mL** is recommended.

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
Stock	powder	Refer to lot	xxx
# 1	Refer to lot	Refer to lot	50 ng/mL
# 2	100 µL of 1	300 µL	12.5 ng/mL
# 3	100 µL of 2	300 µL	3.125 ng/mL
# 4	100 µL of 3	100 µL	0.78 ng/mL

**Positive Control** - Reconstitute the Positive Control with refer to lot Dilution Buffer.

**Detection Antibody** - Reconstitute the Detection Antibody Concentrate with refer to lot of **Dilution Buffer** to produce a 10-fold concentrated stock solution. Allow the concentrated solution to sit for at least 5 minutes until completely dissolved. Pipette 9.45 mL of **Dilution Buffer** into a 15 mL centrifuge tube and transfer 1.05 mL of 10-fold concentrated stock solution to prepare working solution.

**Streptavidin HRP Conjugate** - Pipette 9.395 mL of **Dilution Buffer** into a 15 mL centrifuge tube and transfer 105 µL of 100-fold concentrated stock solution to prepare working solution (**protect from light**). **DO NOT FREEZE.**

### ELISA PROTOCOL

**Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples**

should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents, standard dilutions, positive control and samples as directed previously.
2. Add 100 µL per well of **Dilution Buffer** to Blank wells.
3. Add 100 µL per well of **Standard dilutions, samples, or positive control**. Cover with plate sealer and incubate for 2 hours on microplate shaker at room temperature.
4. Aspirate and wash each well with 300 µL of **1x Wash Buffer** four times. After the last wash, aspirate any remaining 1x Wash Buffer, invert the plate and blot against clean paper towel(s).
5. Add 100 µL per well of **Detection Antibody working solution**. Cover with plate sealer and incubate for 2 hours on microplate shaker at room temperature.
6. Repeat the aspiration and wash as in step 4.
7. Add 100 µL per well of **Streptavidin HRP Conjugate working solution**. Cover with plate sealer and incubate for 60 minutes on microplate shaker at room temperature. **Protect from light.**
8. Repeat the aspiration and wash as in step 4.
9. Add 100 µL per well of **Substrate Solution**. Incubate for refer to lot on microplate shaker at room temperature. **Protect from light.**
10. Add 100 µL per well of **Stop Solution**. The color in the wells should change from blue to yellow. If the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
11. Read plate using a microplate reader set to 450 nm within 3 minutes.

**CALCULATION OF RESULTS**

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

**SPECIFICITY**

Protein	Cross-reactivity (%)
Human Epigen	100

**TYPICAL STANDARD CURVE**

This standard curve is for demonstration only. A new standard curve should be made for each set of samples assayed.

STANDARD (NG/ML)	AVERAGE OD450 (CORRECTED)
Blank	0 (refer to lot)
0.78	0.036
3.125	0.124
12.5	0.457
50	1.386
200 (optional)	2.487

**SUMMARY OF ASSAY PROCEDURE**

