

HUMAN APOLIPOPROTEIN A IV (APOA4) ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF
HUMAN APOA4 CONCENTRATIONS IN SERUM AND
PLASMA



ALWAYS REFER TO LOT SPECIFIC PROTOCOL
PROVIDED WITH EACH KIT FOR
INSTRUCTIONS. PROTOCOL MUST BE
READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN
DIAGNOSTIC PROCEDURES.

PRODUCT INFORMATION:

THIS KIT IS FOR ONE TIME USE ONLY.

| | |
|---|---|
| ELISA NAME | APOA4 (HUMAN) ELISA KIT |
| Catalog No. | SK00401-01 |
| Lot No. | |
| Formulation | 96 T |
| Standard range | 1.95 ~ 500 ng/mL |
| Sensitivity | 0.4 ng/mL |
| Sample Volume | 100 µL |
| Sample Type | Serum, Plasma |
| Dilution Factor | <i>Optimal dilutions should be determined by each laboratory for each application</i> |
| Specificity | Human APOA4 |
| Calibration | Human APOA4 recombinant |
| Intra-assay Precision | 4 - 6% |
| Inter-assay Precision | 8 - 10% |
| Storage | 2 - 8° C for 1 month, see page 2-3 for more information |
| This kit contains sufficient materials to run approximately 35 samples duplicated provided that assay is run according to protocol. | |

ORDER CONTACT:

AVISCERA BIOSCIENCE, INC.

2348 Walsh Ave., Suite C

Santa Clara, CA 95051

USA

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: Sales@AvisceraBioscience.com

Info@AvisceraBioscience.com

www.AvisceraBioscience.net

www.AvisceraBioscience.com

DESCRIPTION

This Human Apolipoprotein A IV (APOA4) ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural APOA4 from serum samples and plasma in a sandwich ELISA format.

This immunoassay contains recombinant human APOA4 and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural human APOA4 samples.

ASSAY OVERVIEW

This assay employs the quantitative sandwich enzyme immunoassay technique. The plate is pre-coated with a monoclonal antibody specific for human APOA4. The capture antibody can bind to the human APOA4 in the standard and samples. After washing the plate of any unbound substances, a biotinylated monoclonal antibody against human APOA4 is added to the wells. After another washing of the plate, Streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution is added to the wells and color develops in direct proportion to the amount of human APOA4 bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

| DESCRIPTION | CODE | QUANTITY |
|--|------------------|-----------------|
| APOA4 Microplate - 96 well polystyrene microplate (12 strips of 8 wells) coated with a monoclonal antibody against human APOA4. | 401-01-01 | 1 plate |
| Human APOA4 Standard – refer to lot of recombinant human APOA4 in a buffered protein base with preservative; lyophilized. | 401-01-02 | 1 vial |
| Detection Antibody Concentrate – 1.05 ml/vial, 10-fold concentrated of biotinylated antibody solution against human APOA4 with preservative | 401-01-03 | 1 vial |
| Streptavidin-HRP Conjugate - 120 µl/vial, 100-fold concentrated solution of Streptavidin conjugate to HRP. | SAHRP | 1 vial |
| Dilution Buffer – 40 mL of buffered protein based solution with preservative. | DB01 | 1 bottle |
| Antibody & HRP Diluent Solution – 25 mL of buffered protein based solution with preservative. | DB08C | 1 bottle |
| Wash Buffer - 50 mL of 10-fold concentrated buffered surfactant, with preservative. | WB01 | 1 bottle |
| TMB Substrate Solution - 11 mL of TMB substrate solution. | TMB01 | 1 bottle |
| Stop Solution - 11 mL of 0.5M HCl solution. | S-STOP | 1 bottle |
| Plate Sealer | EAPS | 1 piece |
| Plastic Pouch | P01 | 1 piece |

STORAGE

Unopened Kit: Store at 2 – 8° C for up to 1 month. For longer storage up to 4 months, unopened Standard, Detection Antibody, Dilution Buffer and Antibody & HRP Diluent Solution should be stored at -20° C. Streptavidin-HRP Conjugate and TMB Substrate Solution should be stored only at 2-8° C. Do not use kit past expiration date.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (400 – 450 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

SAMPLE PREPARATION

Serum or plasma samples may require dilution. **Optimal dilutions should be determined by each laboratory for each application with a sample pretest.**
Use polypropylene test tubes.

REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of 1x Wash Buffer.

APOA4 Standard - Reconstitute the APOA4 standard with refer to lot of Dilution Buffer. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 300 µL of Dilution Buffer into the tube #1 to #5. Use the stock solution to produce a 4-fold dilution series (below). Mix each tube thoroughly before the next transfer. The **500 ng/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 ng/mL).

| TUBE | STANDARD | DILUTION BUFFER | CONCENTRATION |
|-------|--------------|-----------------|---------------|
| stock | powder | Refer to lot | |
| # 1 | Refer to lot | Refer to lot | 500 ng/ml |
| # 2 | 100 µl of 1 | 300 µl | 125 ng/ml |
| # 3 | 100 µl of 2 | 300 µl | 31.25 ng/ml |
| # 4 | 100 µl of 3 | 300 µl | 7.813 ng/ml |
| # 5 | 100 µl of 4 | 300 µl | 1.95 ng/ml |

Detection Antibody - Reconstitute the Detection Antibody Concentrate with 1.05 mL of Antibody & HRP Diluent Solution (DB08C) to prepare 10-fold concentrated stock. Pipette 9.45 mL of Antibody & HRP Diluent Solution (DB08C) into a 15 mL centrifuge tube and transfer 1.05 mL of 10-fold concentrated stock solution to prepare working solution.

Streptavidin-HRP Conjugate - Pipette 11.88 mL of Antibody & HRP Diluent Solution (DB08C) into a 15 mL centrifuge tube and transfer 120 µL of 100-fold concentrated stock solution to prepare working solution (**protect from light**).

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Add 100 µL per well of Dilution Buffer to Blank wells.
3. Add 100 µL of standard dilutions in reverse order of serial dilution, samples, or positive control per well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
4. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with 1x Wash Buffer (300 µL) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
5. Add 100 µL of Detection Antibody working solution to each well. Cover with plate sealer.

Incubate for 2 hours on microplate shaker at room temperature.

6. Repeat the aspiration/wash as in step 4.
7. Add 100 μ L of Streptavidin-HRP Conjugate working solution to each well. Incubate for 60 minutes on microplate shaker at room temperature. **Protect from light.**
8. Repeat the aspiration/wash as in step 4.
9. Add 100 μ L of Substrate Solution to each well. Incubate for refer to lot on microplate shaker at room temperature. **Protect from light.**
10. Add 100 μ L of Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
11. Determine the optical density of each well using a microplate reader set to 450 nm within 3 min.

CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a 4-parameter curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

SPECIFICITY

| PROTEINS | CROSS-REACTIVITY (%) |
|-------------|----------------------|
| Human APOA4 | 100 |
| Human APOH | 0 |

TYPICAL STANDARD CURVE

This standard curve is provided for demonstration only. A new standard curve should be generated for each set of samples assayed.

| STANDARD (NG/ML) | AVERAGE OD450 (CORRECTED) |
|------------------|---------------------------|
| Blank | 0 (0.107) |
| 1.95 | 0.045 |
| 7.8 | 0.231 |
| 31.25 | 0.935 |
| 125 | 2.447 |
| 500 | 3.259 |

SUMMARY OF ASSAY PROCEDURE

PREPARE REAGENTS, SAMPLES AND STANDARDS

Add 100 μ L of standard dilutions, samples, or positive control to the well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 μ L Detection Antibody working solution to each well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 μ L Streptavidin-HRP conjugate working solution to each well. Incubate 60 min on the plate shaker at RT. **Protect from light.**

Aspirate and wash 4 times.

Add 100 μ L Substrate solution to each well. Incubate refer to lot on the plate shaker at RT. **Protect from light.**

Add 100 μ L Stop Solution to each well. Read at 450 nm within 3 min.