

HUMAN TETRANECTIN ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF HUMAN TETRANECTIN CONCENTRATIONS IN EDTA PLASMA



THIS PROTOCOL OR DATA IS PROVIDED FOR DEMONSTRATION ONLY. ALWAYS REFER TO LOT SPECIFIC PROTOCOL PROVIDED WITH EACH KIT FOR INSTRUCTIONS. PROTOCOL MUST BE READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

PURCHASE INFORMATION:
THIS KIT IS FOR ONE TIME USE ONLY.

ELISA NAME	HUMAN TETRANECTIN ELISA
Catalog No.	SK00502-06
Lot No.	
Formulation	96 T
Standard range	1 – 64 ng/mL
Sensitivity	500 pg/mL
Sample Volume	100 µL
Dilution Factor	400-800 fold for EDTA plasma. Optimal dilutions should be determined by each laboratory for each application
Sample Type	EDTA plasma
Specificity	Human Tetranectin only
Calibration	Human Tetranectin recombinant (HEK293)
Intra-assay Precision	4 - 6%
Inter-assay Precision	8 - 12%
Storage	2 - 8° C for 1 month. See page 2~3 for detail
This kit contains sufficient materials to run approximately 35 samples duplicated provided that assay is run according to protocol.	

Order Contact:
AVISCIERA BIOSCIENCE, INC
2348 WALSH AVE., SUITE C
SANTA CLARA, CA 95051
USA
 Tel: 408-982-0300
 Fax: 408-982-0301
 Email: Sales@AvisceraBioscience.com
Info@AvisceraBioscience.com
www.AvisceraBioscience.com
www.AvisceraBioscience.net

DESCRIPTION

This Human Tetranectin/ C-type lectin domain family 3, member B (CLEC3B) ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human Tetranectin from EDTA plasma in a sandwich ELISA format.

This immunoassay contains recombinant human Tetranectin and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural Tetranectin samples.

ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with a monoclonal antibody specific for human Tetranectin. The capture antibody can bind to the human Tetranectin in the standard and samples. After washing the plate of any unbound substances, a biotinylated monoclonal antibody against human Tetranectin is added to the wells. After another washing of the plate, Streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution is added to the wells and color develops in direct proportion to the amount of human Tetranectin bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Each laboratory must determine the optimal dilution factors for the samples being assayed.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
Tetranectin Microplate - 96 well polystyrene microplate (12 strips of 8 wells) coated with a purified antibody against human Tetranectin.	502-06-01	1 plate
Tetranectin Standard – refer to lot of recombinant human Tetranectin in a buffered protein base with preservative; lyophilized.	502-06-02	1 vial
Detection Antibody Concentrate – refer to lot of 10-fold concentrate of biotinylated purified antibody against human Tetranectin with preservative; lyophilized.	502-06-03	1 vial
Positive Control – one vial of recombinant human Tetranectin; lyophilized.	502-06-04	1 vial
Streptavidin-HRP Conjugate – 120 µL/vial of 100-fold concentrated solution of Streptavidin conjugate to HRP.	SAHRP	1 vial
Dilution Buffer – 50 mL of buffered protein based solution with preservative.	DB161	1 bottle
HRP Diluent Solution – 12 mL of buffered protein based solution with preservative.	DB08A	1 bottle
Wash Buffer – 50 mL of 10-fold concentrated buffered surfactant, with preservative.	WB01	1 bottle
TMB Substrate Solution -11 mL of TMB substrate solution.	TMB01	1 bottle
Stop Solution – 11 mL of 0.5M HCl.	S-STOP	1 bottle
Plate Sealer	EAPS	1 piece
Plastic Pouch	P01	1 piece

STORAGE

Unopened Kit: Store at 2 – 8° C for up to 1 month. For longer storage for up to 12 months, unopened Standard, Positive Control, Detection Antibody Concentrate, Dilution Buffer and HRP Diluent

Solution should be stored at -20° C. **Streptavidin-HRP Conjugate** and **TMB Substrate Solution** should be stored only at 2 ~ 8 °C. Do not use kit past expiration date.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

SAMPLE COLLECTION AND STORAGE

Plasma - Collect plasma only using EDTA as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤ -20° C or -70° C. Avoid repeated freeze-thaw cycles.

Special Notice: Serum samples may not be suitable for Tetranectin assay.

SAMPLE PREPARATION

EDTA plasma samples may require a 400-800 dilution. A suggested 20-fold dilution is 10 µL sample + 190 µL of **Dilution Buffer (DB161)**. A final 400-fold dilution is 10 µL of 20-fold diluted sample solution + 190 µL of **Dilution Buffer (DB161)**. A final 800-fold dilution is 5 µL of 20-fold diluted sample solution + 195 µL of **Dilution Buffer (DB161)**.

Optimal dilutions should be determined by each laboratory for each application.
Use polypropylene test tubes.

REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50mL of Wash Buffer Concentrate into deionized or distilled water (450mL) to prepare 500 mL of 1x Wash Buffer.

Tetranectin Standard - Reconstitute the Tetranectin standard with refer to lot of Dilution Buffer. This reconstitution produces a stock solution of 64 ng/mL. Allow the standard to sit for a minimum of 15

minutes with gentle agitation prior to making dilutions. Pipette 250 µL of Dilution Buffer into tubes #1 to #6. Use the stock solution to produce a 2-fold dilution series (below). Mix each tube thoroughly before the next transfer. The **64 ng/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 ng/mL).

Tube	Standard	Dilution Buffer	Concentration
stock	powder	Refer to lot	64 ng/ml
# 1	250 µl of stock	250 µl	32 ng/ml
# 2	250 µl of 1	250µl	16 ng/ml
# 3	250 µl of 2	250µl	8 ng/ml
# 4	250 µl of 3	250µl	4 ng/ml
# 5	250 µl of 4	250µl	2 ng/ml
# 6	250 µl of 4	250µl	1 ng/ml

Positive Control - Reconstitute Positive Control with refer to lot of Dilution Buffer.

Detection Antibody - Reconstitute the Detection Antibody Concentrate with refer to lot of Dilution Buffer to produce a 10-fold concentrated stock solution. Pipette 9.45 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 1.05 mL of 10-fold concentrated stock solution to prepare working solution.

Streptavidin-HRP Conjugate - Pipette 11.88 mL of HRP Diluent Solution (DB08A) into a 15 mL centrifuge tube and transfer 120 µL of 100-fold concentrated stock solution to prepare working solution (*protect from light*).

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. **Blank, standard dilutions, positive control and samples should be assayed in duplicate.** ELISA Protocol may need further optimization.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Add 100 µL per well of **Dilution Buffer** to Blank wells.
3. Add 100 µL of **Standard dilutions** in reverse order of serial dilution, **samples**, or **positive control** per

well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.

4. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with **1x Wash Buffer** (300 µL) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
5. Add 100 µL of **Detection Antibody working solution** to each well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
6. Repeat the aspiration/wash as in step 4.
7. Add 100 µL of **Streptavidin-HRP Conjugate working solution** to each well. Incubate for 1 hour on microplate shaker at room temperature. **Protect from light.**
8. Repeat the aspiration/wash as in step 4.
9. Add 100 µL of **Substrate Solution** to each well. Incubate for refer to lot on microplate shaker at room temperature. **Protect from light.**
10. Add 100 µL of **Stop Solution** to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
11. Determine the optical density of each well using a microplate reader set to 450 nm.

CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

Calculation of samples with a concentration exceeding that of standard 64 ng/mL may result in inaccurate, low human Tetranectin levels. Such samples require further external predilution according to expected human tetranectin values with Dilution Buffer in order to precisely quantify the actual human Tetranectin level.

TYPICAL DATA

This standard curve is provided for demonstration only. A new standard curve should be generated for each set of samples assayed.

Standard (ng/mL)	Average OD450nm (Corrected)
Blank	0 (refer to lot)
1	0.050
2	0.099
4	0.211
8	0.394
16	0.419
32	1.110
64	2.112

SPECIFICITY

PROTEINS	CROSS-REACTIVITY
Human Tetranectin	100%
Human Periostin	0
Human HE4	0

SUMMARY OF ASSAY PROCEDURE

