

HUMAN NEUROTROPHIN 3 (NT3) ELISA KIT

FOR THE QUANTITATIVE DETERMINATION
OF HUMAN NEUROTROPHIN 3
CONCENTRATIONS IN CELL CULTURE
SUPERNATES AND TISSUE HOMOGENATES,
SERUM AND PLASMA



ALWAYS REFER TO LOT SPECIFIC PROTOCOL
PROVIDED WITH EACH KIT FOR
INSTRUCTIONS. PROTOCOL MUST BE
READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN
DIAGNOSTIC PROCEDURES.

PRODUCT INFORMATION:

ELISA NAME	HUMAN NEUROTROPHIN 3 ELISA
Catalog No.	SK00728-06
Lot No.	
Formulation	96 T
Standard Range	31.25 – 2000 pg/mL
Sensitivity	10 pg/mL
Sample Volume	100 µL
Dilution Factor	Optimal dilutions should be determined by each laboratory for each application
Sample Type	Serum and Plasma
Specificity	Human Neurotrophin 3
Calibration	Human Neurotrophin 3 Recombinant
Intra-assay Precision	4 - 6%
Inter-assay Precision	8 - 12%
Storage	2 - 8° C
This kit contains sufficient materials to run 35 samples duplicated provided that assay is run according to protocol.	

Order Contact:

AVISCERA BIOSCIENCE, INC.

2348 Walsh Ave., Suite C

Santa Clara, CA 95051

USA

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: Info@AvisceraBioscience.com

Sales@AvisceraBioscience.com

www.AvisceraBioscience.com

DESCRIPTION

This Human Neurotrophin 3 ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human Neurotrophin 3 from serum and plasma in a sandwich ELISA format.

This immunoassay contains recombinant human Neurotrophin 3 and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural Neurotrophin 3 samples.

ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with an antibody specific for human Neurotrophin 3. The capture antibody can bind to the human Neurotrophin 3 in the standard and samples. After washing the plate of any unbound substances, a biotinylated antibody against human Neurotrophin 3 is added to the wells. After another washing of the plate, Streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution (TMB) is added to the wells and color develops in direct proportion to the amount of human Neurotrophin 3 bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
Neurotrophin 3 Microplate - 96 well polystyrene microplate (12 strips of 8 wells) coated with an antibody against Neurotrophin 3.	728-06-01	1 plate
Neurotrophin 3 Standard – 2000 pg/vial of recombinant Neurotrophin 3 in a buffered protein base with preservative; lyophilized.	728-06-02	1 vial
Detection Antibody Concentrate – 1.05 mL/vial, 10-fold concentrate of biotinylated antibody against human Neurotrophin 3 with preservative; lyophilized.	728-06-03	1 vial
Positive Control – one vial of recombinant human Neurotrophin 3 ; lyophilized.	728-06-04	1 vial
Streptavidin-HRP Conjugate – 120 µL/vial of 100-fold concentrated solution of Streptavidin-HRP with preservative.	SAHRP	1 vial
Dilution Buffer – 60 mL of buffered protein based solution with preservative.	DB01	1 bottle
Wash Buffer - 50 mL of 10-fold concentrated buffered surfactant, with preservative.	WB01	1 bottle
TMB Substrate Solution - 11 mL of TMB Substrate Solution.	TMB01	1 bottle
Stop Solution – 11 mL of 0.5M HCl.	S-STOP	1 bottle
Plate Sealer	EAPS	1 piece
Plastic Pouch	P01	1 piece

STORAGE

Unopened Kit: Store at 2 – 8° C for up to 8 months. For longer storage, unopened Standard, Positive Control and Detection Antibody Concentrate should

be stored at -20° C or -70° C. Do not use kit past expiration date.

Opened / Reconstituted Reagents: Reconstituted Standard (stock) solution and Detection Antibody concentrated solution SHOULD BE STORED at -20° C or -70° C for up to one month. Streptavidin-HRP Conjugate 100-fold concentrated solution and TMB Substrate Solution can be stored at 2 – 8° C for up to 8 months (**DO NOT FREEZE and PROTECT FROM LIGHT**). All other components may be stored at 2 – 8° C for up to 8 months.

Microplate Wells: Return unused strips to the plastic pouch with the desiccant pack. Microplate may be stored for up to 6 months at 2 – 8° C after opening.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

SAMPLE PREPARATION

Optimal dilutions should be determined by each laboratory for each application with a pretest.
Use polypropylene test tubes.

REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of 1x Wash Buffer.

Neurotrophin 3 Standard - Reconstitute the Neurotrophin 3 standard with 1.0 mL of Dilution Buffer. This reconstitution produces a stock solution of 2000 pg/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 250 µL of Dilution Buffer into tubes #1 to #6. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The **2000 pg/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 ng/mL).

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
stock	Powder	1.0 ml	2000 pg/ml
# 1	250µl of stock	250µl	1000 pg/ml
# 2	250µl of 1	250µl	500 pg/ml
# 3	250µl of 2	250µl	250 pg/ml
# 4	250µl of 3	250µl	125 pg/ml
# 5	250µl of 4	250µl	62.5 pg/ml
# 6	250µl of 5	250µl	31.25 pg/ml

Positive Control - Reconstitute the Positive Control with refer to lot specific label of Dilution Buffer.

Note: Positive Control solution could be reused within a few days if stored at -20° C or -70° C.

Detection Antibody Concentrate – Reconstitute the Detection Antibody Concentrate with 1.05 mL of Dilution Buffer to prepare a 10-fold concentrated solution. Pipette 9.45 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer the 1.05 mL of 10-fold concentrated solution to the tube to make 1x working solution.

Streptavidin-HRP Conjugate - Pipette 11.88 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 120 µL of 100-fold concentrated stock solution to prepare working solution. **Note:** 1x working solution of Streptavidin-HRP Conjugate should be used within a few days (**protect from light**).

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess microplate strips from the plate frame, return them to the plastic pouch with the desiccant pack.
3. Add 100µL per well of Dilution Buffer to Blank wells.
4. Add 100µL of standard dilutions in reverse order of serial dilution, samples, or positive per well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
5. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with 1x Wash Buffer (300µL) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential for good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
6. Add 100µL of Detection Antibody working solution to each well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
7. Repeat the aspiration/wash as in step 5.
8. Add 100µL of Streptavidin-HRP Conjugate working solution to each well. Incubate for 1 hour on microplate shaker at room temperature. **Protect from light.**
9. Repeat the aspiration/wash as in step 5.
10. Add 100µL of Substrate Solution to each well. Incubate 8-12 minutes on microplate shaker at room temperature. **Protect from light.**
11. Add 100µL of Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Determine the optical density of each well within 15 minutes, using a microplate reader set to 450nm.

CALCULATION OF RESULTS

Average the duplicate readings for each standard, positive control and sample, and subtract the average zero standard optical density. Create a standard curve by reducing the data using computer software capable of generating a log-log curve fit. As

an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the human Neurotrophin 3 concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

Calculation of samples with a concentration exceeding that of standard 2000pg/mL may result in inaccurate, low human Neurotrophin 3 levels. Such samples require further external pre-dilution according to expected human Neurotrophin 3 values with Dilution Buffer in order to precisely quantify the actual human Neurotrophin 3 level.

TYPICAL DATA









This standard curve is provided for demonstration only. A new standard curve should be generated for each set of samples assayed.

STANDARD (NG/ML)	CORRECTED (450NM)
Blank	0 (0.098)
31.25	0.095
62.5	0.221
125	0.401
250	0.632
500	1.063
1000	1.672
2000	2.094

SPECIFICITY

PROTEIN NAME	CROSS-REACTIVITY
Human Neurotrophin 3	100%
Human Neurotrophin 4	0
Human BDNF	0
Human Beta NGF	0

SUMMARY OF ASSAY PROCEDURE

PREPARE REAGENTS, SAMPLES AND STANDARDS

Add 100 µl of standard dilutions, samples, or positive control to the well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µl Detection Antibody working solution to each well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µl Streptavidin-HRP conjugate working solution to each well. Incubate 1 hour on plate shaker at RT. Protect from light.

Aspirate and wash 4 times.

Add 100 µl Substrate Solution to each well. Incubate 8-12 min on plate shaker at RT. Protect from light.

Add 100 µl Stop Solution to each well. Read at 450nm within 15 min.