INTERLUKIN-6 (IL-6) (HUMAN) ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF HUMAN INTERLUKIN-6 CONCENTRATIONS IN RECOMBINANT PROTEIN SAMPLES



ALWAYS REFER TO LOT SPECIFIC PROTCOL PROVIDED WITH EACH KIT FOR INSTRUCTIONS. PROTOCOL MUST BE READ AND CHECK ALL ITEMS OF EACH KIT BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY.NOT FOR USE IN DIAGNOSTIC PROCEDURES.

PRODUCT INFORMATION:

THIS IS FOR ONE TIME USE ONLY.

ELISA NAME	IL-6 (HUMAN) ELISA KIT	
Catalog No.	SK00110-01	
Lot No.:		
Formulation	96 T	
Standard range	3.125 – 200 pg/mL	
Sensitivity	1.5 pg/mL	
Sample Volume	100 μL	
Dilution Factor	Optimal dilutions should be determined by each laboratory for each application	
Sample Type	Recombinant Protein	
Specificity	Human	
Calibration	Human IL-6 recombinant	
Intra-assay Precision	4 - 6%	
Inter-assay Precision	8 - 10%	
Storage	2 – 8 °C for 1 month, see page 2-3 for detail	
This kit contains sufficient materials to run 35~40 samples duplicated provided that assay is run according to protocol.		

THIS IS ONE TIME USE ONLY

Order Contact:

AVISCERA BIOSCIENCE, INC.

2348 Walsh Ave., Suite C

Santa Clara, CA 95051

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: Sales@AvisceraBioscience.com

Info@AvisceraBioscience.com

www.AvisceraBioscience.com www.AvisceraBioscience.net

DESCRIPTION

This Human Interlukin-6 (IL-6) ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human IL-6 from recombinant protein samples or cell cultures in a sandwich ELISA format.

This immunoassay contains recombinant human IL-6 and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural IL-6 samples. If testing samples are human serum or plasma, please order High Sensitivity IL-6 (Human) ELISA Kit SK00110-01A.

ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with a monoclonal antibody specific for human IL-6. The capture antibody can bind to the human IL-6 in the standard and samples. After washing the plate of any unbound substances, a biotinylated monoclonal antibody against human IL-6 is added to the wells. After another washing of the plate, the Streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution (TMB) is added to the wells and color develops in direct proportion to the amount of IL-6 bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
IL-6 Microplate – 96 well microplate coated with monoclonal antibody	110-01-01	1 plate
specific for human IL-6.		
IL-6 Standard – 400 pg/ vial of lyophilized recombinant human IL-6.	110-01-02	1 vial
Detection Antibody		
Concentrate – 1.05 mL/vial of 10-fold concentrate of lyophilized	110-01-03	1 vial
biotinylated monoclonal antibody against human IL- 6.		
Streptavidin-HRP Conjugate – 120 μL/vial of 100-fold concentrated solution of Streptavidin- HRP conjugate.	SAHRP	1 vial
Dilution Buffer – 40 mL of buffered solution with preservative.	DB03A	1 bottle
HRP Diluent Solution – 12 mL of buffered solution with preservative.	DB08C	1 bottle
Wash Buffer – 50 mL of 10-fold concentrated buffered surfactant with preservative.	WB01	1 bottle
TMB Substrate Solution – 11 mL of TMB substrate solution.	TMB01	1 bottle
Stop Solution – 11 mL of 0.5M HCl.	S-STOP	1 bottle
Plate Sealer	EAPS	1 piece
Plastic Pouch	P01	1 piece

STORAGE

Unopened Kit: Store at 2-8 °C for up to 1 month. For longer storage for up to 10 months, unopened Standard, Detection Antibody Concentrate, Dilution Buffer and HRP Diluent Solution should be stored at $-20 \sim -70$ °C. Streptavidin-HRP Conjugate 100-fold concentrated solution and TMB Substrate Solution should be stored only at 2-8 °C (**DO NOT FREEZE** and **PROTECT FROM LIGHT**). Do not use kit past expiration date.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (350 400 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

SAMPLE COLLECTION AND STORAGE

Cell Culture- All samples should be used animal free media. The fetal bovine serum or other animal serum may crossreact and or interface with this elisa assay.

If testing samples are human serum or plasma, please contact with Aviscera Bioscience and order SK00110-01A, High Sensitivity II-6 (Human) ELISA Kit.

Note: Use Aprotinin (enzyme inhibitor) (Code No.: 00700-01-25) for ALL sample collection to prevent sample degradation. 0.5 TIU per ml of sample solution.

SAMPLE PREPARATION

Optimal dilutions should be determined by each laboratory for each application with a pretest. Use polypropylene test tubes.

REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer – Dilute 50 mL of Wash Buffer Concentrate into 450 mL distilled or deionized water to make 500 mL of 1x Wash Buffer. If crystals have formed in the concentrate, warm bottle in a water bath until the crystals have completely dissolved.

IL-6 Standard – Reconstitute the human IL-6 standard with 1ml of Dilution Buffer. The concentration of the reconstituted stock solution is 400 pg/mL. Allow the stock standard to sit for at least 15 minutes with gentle agitation until completely dissolved prior to making standard dilutions (see below). Mix each tube thoroughly before the next transfer.

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
- I			100 / 1
Stock	powder	1 ml	400 pg/ml
#1	250 μL of stock	250 μL	200 pg/mL
# 2	250μL of 1	250 μL	100 pg/mL
#3	250μL of 2	250 μL	50 pg/mL
# 4	250μL of 3	250 μL	255 pg/mL
# 5	250μL of 4	250 μL	12.5 pg/mL
# 6	250μL of 5	250 μL	6.25 pg/mL
#7	250μL of 6	250 μL	3.125 pg/mL

Detection Antibody - Reconstitute the Detection Antibody Concentrate with 1.05 mL of Dilution Buffer to produce a 10-fold concentrated stock solution. Allow the concentrated solution to sit for at least 5 minutes until completely dissolved. Pipette 9.45 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 1.05 mL of 10-fold concentrated stock solution to prepare working solution.

Streptavidin-HRP Conjugate - Pipette 11.88 mL of HRP Diluent Solution into a 15 mL centrifuge tube and transfer 120 μ L of 100-fold concentrated stock solution to prepare working solution. **Note:** (protect from light). DO NOT FREEZE. The working solution of Streptavidin-HRP Conjugate should be freshly prepared and used within a few hours.

Positive Control - Reconstitute the Positive Control refer to lot.

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

- 1. Prepare all reagents, standard dilutions, positive control and samples as directed previously.
- 2. Remove unneeded microplate strips from the plate frame and return them to the plastic pouch with the desiccant pack.
- 3. Add 100 µL per well of **Dilution Buffer** to Blank wells.
- 4. Add 100 μL per well of Standard Dilutions in reverse order of serial dilution, sample, or positive control. Cover with plate sealer and incubate for 2 hours on microplate shaker at room temperature. (Please see plate layout provided.)
- 5. Aspirate and wash each well with 300 μL of 1x Wash Buffer four times. After the last wash, aspirate any remaining 1x Wash Buffer, invert the plate and blot against clean paper towel(s).
- 6. Add 100 μ L per well of **Detection Antibody** working solution. Cover with plate sealer and incubate for 2 hours on microplate shaker at room temperature.
- 7. Repeat the aspiration and wash as in step 5.
- Add 100 μL per well of working solution of Streptavidin HRP conjugate. Cover with plate sealer and incubate for 60 minutes on microplate shaker at room temperature. Protect from light.
- 9. Repeat the aspiration and wash as in step 5.
- 10. Add 100 μ L per well of **Substrate Solution**. Incubate for refer to lot on microplate shaker at room temperature. **Protect from light.**
- 11. Add 100 μ L per well of **Stop Solution**. The color in the wells should change from blue to yellow. If the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
- 12. Read plate using a microplate reader set to 450 nm within 5 minutes.

SPECIFICITY

PROTEIN	CROSS-REACTIVITY
Human IL-6	100%
Human IL-4	0
Human IL-10	0
Human IL-15	0
Human IL-33	0

TYPICAL STANDARD CURVE

This standard curve is for demonstration only. A new standard curve should be made for each set of samples assayed.

STANDARD (PG/ML)	AVERAGE OD450 NM (CORRECTED)
Blank	0 (0.101)
3.125	0.031
6.25	0.059
12.5	0.097
25	0.178
50	0.279
100	0.699
200	1.239

CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log or 4-parameter curve fit to more accurately quantify the standard dilutions. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

SUMMARY OF ASSAY PROCEDURE

PREPARE REAGENTS, SAMPLES AND STANDARD DILUTIONS



Add 100 μL of standard dilutions, samples and positive control. Cover with plate sealer and incubate 2 hours on microplate shaker at RT.



Aspirate and wash 4 times.



Add 100 μ L per well of Detection Antibody working solution. Cover with plate sealer and incubate 2 hours on microplate shaker at RT.



Aspirate and wash 4 times.



Add 100 μ L per well of Streptavidin-HRP Conjugate working solution. Cover with plate sealer and incubate 60 minutes on microplate shaker at RT. **Protect from light.**



Aspirate and wash 4 times.



Add 100 μL per well of Substrate Solution. Incubate refer to lot on microplate shaker at RT. **Protect from light.**



Add 100 μL per well of Stop Solution. Read at 450 nm within 5 minutes.