

HUMAN SOLUBLE CD44 ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF
HUMAN SOLUBLE CD44 CONCENTRATIONS IN
SERUM AND PLASMA



ALWAYS REFER TO LOT SPECIFIC PROTOCOL
PROVIDED WITH EACH KIT FOR
INSTRUCTIONS. PROTOCOL MUST BE
READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN
DIAGNOSTIC PROCEDURES.

PRODUCT INFORMATION:

ELISA NAME	HUMAN SOLUBLE CD44 ELISA
Catalog No.	SK00481-01
Formulation	96 T
Lot No.	
Standard range	125 - 8000 pg/mL
Sensitivity	50 pg/mL
Sample Volume	100 μ L
Dilution Factor	40 (Optimal dilutions should be determined by each laboratory for each application)
Sample Type	Serum and EDTA Plasma
Specificity	Human soluble CD44
Calibration	Human Soluble CD44 recombinant
Intra-assay Precision	4 - 6%
Inter-assay Precision	8 - 10%
Storage	2 – 8° C
This kit contains sufficient materials to run 35 samples duplicated provided that assay is run according to protocol.	

Order Contact:

AVISCERA BIOSCIENCE, INC.

2348 Walsh Ave., Suite C

Santa Clara, CA 95051

USA

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: Sales@AvisceraBioscience.com

Info@AvisceraBioscience.com

www.AvisceraBioscience.com

DESCRIPTION

This human Soluble CD44 ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human soluble CD44 (sCD44) from serum and plasma in a sandwich ELISA format.

This immunoassay contains recombinant sCD44 and antibody raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural sCD44 samples.

ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with an antibody specific for human sCD44. The capture antibody can bind to the human sCD44 in the standard and samples. After washing the plate of any unbound substances, an antibody conjugated to horseradish-peroxidase (HRP) against human sCD44 is added to the wells. After the last wash to remove any unbound enzyme, a TMB substrate solution is added to the wells and color develops in direct proportion to the amount of human sCD44 bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
sCD44 Microplate – 96 well microplate coated with an antibody specific for sCD44.	481-01-01	1 plate
sCD44 Standard – 8000 pg/vial of lyophilized recombinant sCD44.	481-01-02	1 vial
Detection Antibody-HRP Conjugate – 110 µL/vial of 100-fold concentrated solution of antibody conjugated to HRP against sCD44.	481-01-03	1 vial
Positive Control – one vial of lyophilized recombinant sCD44.	481-01-04	1 vial
Dilution Buffer – 60 mL of buffered solution with preservative.	DB04	1 bottle
Detection Antibody-HRP Diluent Solution – 12 mL of buffered solution with preservative.	DB10	1 bottle
Wash Buffer – 50 mL of 10-fold concentrated buffered surfactant with preservative.	WB01	1 bottle
TMB Substrate Solution – 11 mL of TMB substrate solution.	TMB01	1 bottle
Stop Solution – 11 mL of 0.5M HCl.	S-STOP	1 bottle
Plate Sealer	EAPS	1 piece
Plastic Pouch	P01	1 piece

STORAGE

Unopened Kit: Store at 2 – 8° C for up to 6 months. For longer storage, unopened Standard and Positive Control should be stored at -20° C or -70° C. Detection antibody-HRP Conjugate 100-fold concentrated solution should be stored at 2 – 8° C (protect from light). TMB Substrate Solution can be stored at 2 – 8° C for up to 6 months (**DO NOT FREEZE** and **PROTECT FROM LIGHT**). All other components can be stored at 2 – 8° C for up to 6 months. Do not use kit past expiration date.

Opened / Reconstituted Reagents: Reconstituted Standard (stock) solution SHOULD BE STORED at -20° C or -70° C for up to one month.

Microplate Wells: Return unused strips to the plastic pouch with the desiccant pack. Microplate may be stored for up to 6 months at 2 – 8° C.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

SAMPLE COLLECTION AND STORAGE

Serum – Use a serum separator tube (SST). Allow blood to clot for 30 minutes. Centrifuge at 1000 x g for 15 minutes and collect serum. Assay samples immediately or aliquot and store at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Plasma – Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge at 1000 x g for 15 minutes and collect plasma. Assay samples immediately or aliquot and store at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Optional: Use Aprotinin (enzyme inhibitor) for ALL sample collection to prevent sample degradation. 0.5 TIU per mL of sample solution.

SAMPLE PREPARATION

Serum and Plasma samples may need a 40-fold or greater dilution. A suggested 40-fold dilution is 6 μL sample + 234 μL Dilution Buffer.

Optimal dilutions should be determined by each laboratory for each application with a pretest. Use polypropylene test tubes.

REAGENT PREPARATION


Bring all reagents to room temperature before use.

Wash Buffer – If crystals have formed in the concentrate, warm bottle in a water bath until the crystals have completely dissolved. Dilute 50 mL of

Wash Buffer Concentrate into 450 mL distilled or deionized water to make 500 mL of 1x Wash Buffer.

sCD44 Standard – Reconstitute the sCD44 standard with 1.0 mL of Dilution Buffer. The concentration of the reconstituted stock solution is 8000 pg/mL. Allow the stock standard to sit for at least 15 minutes with gentle agitation until completely dissolved prior to making standard dilutions (see below). Mix each tube thoroughly before the next transfer.

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
Stock	powder	1.0 mL	8000 pg/mL
# 1	250 μL of stock	250 μL	4000 pg/mL
# 2	250 μL of 1	250 μL	2000 pg/mL
# 3	250 μL of 2	250 μL	1000 pg/mL
# 4	250 μL of 3	250 μL	500 pg/mL
# 5	250 μL of 4	250 μL	250 pg/mL
# 6	250 μL of 5	250 μL	125 pg/mL



Positive Control - Reconstitute the Positive Control with 0.5 mL Dilution Buffer. **Note:** Positive Control could be used within a few days if stored at -20°C or -70°C .

Detection Antibody-HRP Conjugate - Pipette 10.89 mL of **Detection Antibody-HRP Diluent Solution (DB10)** into a 15 mL centrifuge tube and transfer 110 μL of 100-fold concentrated stock solution to prepare working solution. **Note:** 1x working solution of Detection Antibody-HRP should be used within a few days (**protect from light**). **DO NOT FREEZE.**

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank,

standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents, standard dilutions, positive control and samples as directed previously.
2. Remove unneeded microplate strips from the plate frame and return them to the plastic pouch with the desiccant pack.
3. Add 100 μ L per well of **Dilution Buffer** to Blank wells.
4. Add 100 μ L per well of **Standard Dilutions** in reverse order of serial dilution from #6-S, **sample**, or **positive control**. Cover with plate sealer and incubate for 2 hours on microplate shaker at room temperature.
5. Aspirate and wash each well with 300 μ L of **1x Wash Buffer** four times. After the last wash, aspirate any remaining 1x Wash Buffer, invert the plate and blot against clean paper towel(s).
6. Add 100 μ L per well of **Detection Antibody-HRP Conjugate working solution**. Cover with plate sealer and incubate for 1 hour on microplate shaker at room temperature. **Protect from light.**
7. Repeat the aspiration and wash as in step 5.
8. Add 100 μ L per well of **Substrate Solution**. Incubate for 25-35 minutes on microplate shaker at room temperature. **Protect from light.**
9. Add 100 μ L per well of **Stop Solution**. The color in the wells should change from blue to yellow. If the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
10. Read plate using a microplate reader set to 450 nm within 15 minutes.

CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

SPECIFICITY

Protein	Cross-reactivity
Human sCD44	100%
Human Gas6	0
Human Gas1	0







TYPICAL STANDARD CURVE

This standard curve is for demonstration only. A new standard curve should be made for each set of samples assayed.

sCD44 STANDARD (PG/ML)	AVERAGE OD450 (CORRECTED)
Blank	0 (0.058)
125	0.015
250	0.038
500	0.066
1000	0.141
2000	0.320
4000	0.868
8000	1.715

- Lot No.:
- Positive Control:

SUMMARY OF ASSAY PROCEDURE

PREPARE REAGENTS, SAMPLES AND STANDARD DILUTIONS

Add 100 μ L of standard dilutions, samples and positive control. Cover with plate sealer and incubate 2 hours on microplate shaker at RT.

Aspirate and wash 4 times.

Add 100 μ L per well of Detection Antibody-HRP conjugate working solution. Cover with plate sealer and incubate 1 hour on microplate shaker at RT. Protect from light.

Aspirate and wash 4 times.

Add 100 μ L per well of Substrate Solution. Incubate 25-35 min on microplate shaker at RT. Protect from light.

Add 100 μ L per well of Stop Solution. Read at 450 nm within 15 minutes.